

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ052523\
 Data File : PQ061235.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 May 2023 16:02
 Operator : YP\AJ
 Sample : 02921-06
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 PCB-051223-103-013

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 25 21:29:19 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ052523.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu May 25 13:25:08 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.405	2.761	93484417	54386808	18.657	18.826
2) SA Decachlor...	8.584	7.535	75094885	70839877	19.414	19.454

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ052523\
Data File : PQ061235.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 May 2023 16:02
Operator : YP\AJ
Sample : 02921-06
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
PCB-051223-103-013

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 25 21:29:19 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ052523.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu May 25 13:25:08 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.405 min
Delta R.T.: 0.000 min
Response: 93484417
Conc: 18.66 ng/ml

Instrument :
ECD_Q
ClientSampleId :
PCB-051223-103-013




#1 Tetrachloro-m-xylene

R.T.: 2.761 min
Delta R.T.: 0.001 min
Response: 54386808
Conc: 18.83 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.584 min
Delta R.T.: -0.001 min
Response: 75094885
Conc: 19.41 ng/ml



#2 Decachlorobiphenyl

R.T.: 7.535 min
Delta R.T.: 0.000 min
Response: 70839877
Conc: 19.45 ng/ml