

284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CASE NARRATIVE

Weston Solutions, Inc. Project Name: RFP 792

Project # N/A

Chemtech Project # N2873 Test Name: Pesticide-TCL

A. Number of Samples and Date of Receipt:

14 Solid samples were received on 05/14/2022.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: Cyanide, Mercury, Metals ICP-TAL, METALS TAL+CN, PCB, Pesticide-TCL and SVOC-TCL BNA -20. This data package contains results for Pesticide-TCL.

C. Analytical Techniques:

The analyses were performed on instrument GCECD_D. The front column is ZB-MR2 which is 30 meters, 0.32 mm ID, 0.25 um df, Catalog #: 7HM-G017-11. The rear column is ZBMR1 which is 30 meters, 0.32 mm ID, 0.5 um df, Catalog # 7HM-G016-17. The analysis of Pesticide-TCLs was based on method 8081B and extraction was done based on method 3541.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries met the acceptable criteria.

The Retention Times were acceptable for all samples.

The MS recoveries met the requirements for all compounds.

The MSD recoveries met the acceptable requirements.

The RPD met criteria.

The Blank Spike met requirements for all samples.

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements.

The Continuous Calibration File ID PD069790.D met the requirements except for Heptachlor epoxide is failing in 2nd column. However %deviation is within QC limit for 1st column. Also closing CCC is passing in both columns for all compounds.

E. Additional Comments:

The soil samples results are based on a dry weight basis.

F. Calculation for the Concentration in Soil Samples

Concentration ug/Kg (Dry weight basis) = $\underline{(Ax) (Vt) (DF) (GPC)}$ (CF) (Vi) (Ws) (D) 284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

Where,

Ax = Response (peak area or height) of the compound to be measured.

CF = Mean Calibration Factor from the initial calibration (area/ng).

Vt = Volume of the concentrated extract in uL

Vi = Volume of extract injected (uL). (If a single injection is made onto two columns, use ½ the volume in the syringe as the volume injected onto each column).

Ws = Weight of sample extracted (g).

D = % dry weight or <u>100 - % Moisture</u>

100

 $GPC = \underline{Vin} = GPC$ factor (If no GPC is performed, GPC=1)

vout

DF = Dilution Factor.

G. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_			