

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ052722\
 Data File : PQ057753.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 May 2022 16:20
 Operator : YP\AJ
 Sample : AR1660ICC800
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR16604009

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 28 00:58:04 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ052722CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat May 28 00:48:23 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.690	4.053	556.3E6	456.3E6	40.375	40.160
2)	SA Decachlor...	10.616	9.214	917.4E6	683.6E6	81.491	80.985
Target Compounds							
3)	L1 AR-1016-1	5.875	5.162	271.7E6	287.5E6	789.555	800.208
4)	L1 AR-1016-2	5.898	5.182	474.6E6	414.0E6	819.411	802.216
5)	L1 AR-1016-3	5.960	5.363	277.9E6	209.1E6	808.641	800.685
6)	L1 AR-1016-4	6.061	5.401	241.4E6	161.1E6	807.270	801.931
7)	L1 AR-1016-5	6.355	5.620	199.5E6	210.1E6	804.673	798.817
31)	L7 AR-1260-1	7.483	6.662	306.6E6	374.6E6	815.337	807.154
32)	L7 AR-1260-2	7.740	6.847	380.8E6	432.6E6	809.296	794.164
33)	L7 AR-1260-3	8.027	7.006	483.5E6	430.7E6	815.218	801.298
34)	L7 AR-1260-4	8.340	7.480	358.8E6	357.4E6	817.813	811.868
35)	L7 AR-1260-5	8.681	7.717	823.1E6	879.2E6	814.207	810.273

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ052722\
Data File : PQ057753.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 May 2022 16:20
Operator : YP\AJ
Sample : AR1660ICC800
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR16604009

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 28 00:58:04 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ052722CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat May 28 00:48:23 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

