

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ052722\
 Data File : PQ057769.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 May 2022 20:13
 Operator : YP\AJ
 Sample : AR1242ICC1600
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR12425025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 28 02:13:09 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ052722CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat May 28 02:09:30 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.689	4.054	1128.1E6	937.7E6	76.479	78.568
2) SA Decachlor...	10.614	9.212	1807.2E6	1401.5E6	152.177	153.941
Target Compounds						
16) L4 AR-1242-1	5.875	5.161	435.5E6	454.2E6	1512.630	1533.028
17) L4 AR-1242-2	5.897	5.181	738.2E6	660.4E6	1504.569	1541.600
18) L4 AR-1242-3	5.960	5.362	443.9E6	334.5E6	1507.013	1539.385
19) L4 AR-1242-4	6.062	5.444	388.3E6	312.7E6	1516.052	1515.710
20) L4 AR-1242-5	6.800	5.973	366.0E6	408.4E6	1506.285	1530.693

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ052722\
Data File : PQ057769.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 May 2022 20:13
Operator : YP\AJ
Sample : AR1242ICC1600
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR12425025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 28 02:13:09 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ052722CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat May 28 02:09:30 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

