Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ052819\
Data File : PQ040293.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 May 2019 09:01
Operator : SM\AJ
Sample : AR1254CCC400
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1254341

Manual Integrations APPROVED
Ankita
5/29/2019 8:54:52 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 28 14:19:49 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ052219CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu May 23 04:10:58 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	5.599	4.735	53748352	39549835	16.407	18.414
2)	SA Decachlor...	11.910	10.456	259.7E6	157.4E6	36.352m	37.865
	Target Compounds						
26)	L6 AR-1254-1	8.005	7.129	55849872	61945183	320.251	323.413
27)	L6 AR-1254-2	8.242	7.296	85886435	54496619	310.560	313.685
28)	L6 AR-1254-3	8.636	7.743	95395333	92401646	321.430	319.748
29)	L6 AR-1254-4	8.939	7.994	66208362	53613656	318.572	300.276m
30)	L6 AR-1254-5	9.377	8.438	95190935	86263581	330.531	334.835

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ052819\
Data File : PQ040293.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 May 2019 09:01
Operator : SM\AJ
Sample : AR1254CCC400
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1254341

Manual Integrations APPROVED
Ankita
5/29/2019 8:54:52 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 28 14:19:49 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ052219CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu May 23 04:10:58 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm