

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ052821\
 Data File : PQ053731.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 May 2021 17:20
 Operator : DD\AJ
 Sample : AR1660CCC400
 Misc :
 ALS Vial : 3 Sample Multiplier: 1


Instrument :
 ECD_Q
 LabSampleId :
 AR1660CCC400

Manual Integrations
 APPROVED

Ankita
 6/1/2021 1:45:17 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 31 02:55:00 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ052021CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu May 20 06:23:48 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(31) AR-1260-1 (L7)
 R.T. Response Conc
 8.23 442463442 457.70
 8.49 583316062 461.10
 8.86 466537790 470.95
 9.10 457342000 475.37
 9.45 983665891 433.28

(31) AR-1260-1 #2 (L7)
 R.T. Response Conc
 7.08 320433157 479.86
 7.27 437796097 467.01
 7.43 349532509 445.53
 7.91 306949245 474.79
 8.16 841517023 474.07

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ052821\
Data File : PQ053731.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 May 2021 17:20
Operator : DD\AJ
Sample : AR1660CCC400
Misc :
ALS Vial : 3 Sample Multiplier: 1


Instrument :
ECD_Q
LabSampleId :
AR1660CCC400

Manual Integrations
APPROVED

Ankita
6/1/2021 1:45:17 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 31 02:55:00 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ052021CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu May 20 06:23:48 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(31) AR-1260-1 #2 (L7)
R.T. Response Conc
8.23 395945558 409.58
8.49 526184152 415.94
8.86 433491355 437.59
9.10 429430077 446.35
9.45 983665891 433.28

(31) AR-1260-1 #2 (L7)
R.T. Response Conc
7.08 292406330 437.89
7.27 421522085 449.65
7.43 349532509 445.53
7.91 306949245 474.79
8.16 841517023 474.07

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ052821\
 Data File : PQ053731.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 May 2021 17:20
 Operator : DD\AJ
 Sample : AR1660CCC400
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_Q
 LabSampleId :
 AR1660CCC400

Manual Integrations
 APPROVED

Ankita
 6/1/2021 1:45:17 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 31 02:55:00 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ052021CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu May 20 06:23:48 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	5.217	4.232	449.5E6	244.3E6	19.822	22.682
2) SA Decachlor...	11.400	9.567	884.1E6	571.7E6	39.276	51.776 #
Target Compounds						
3) L1 AR-1016-1	6.542	5.491	321.2E6	184.9E6	389.178	435.042
4) L1 AR-1016-2	6.567	5.511	515.7E6	279.8E6	404.723	445.475
5) L1 AR-1016-3	6.633	5.704	326.1E6	142.4E6	423.961	443.030
6) L1 AR-1016-4	6.741	5.753	260.2E6	111.4E6	408.756	445.256
7) L1 AR-1016-5	7.054	5.985	236.9E6	97135559	399.689	297.601 #
31) L7 AR-1260-1	8.227	7.077	395.9E6	292.4E6	409.579m	437.892m
32) L7 AR-1260-2	8.491	7.274	526.2E6	421.5E6	415.938m	449.654m
33) L7 AR-1260-3	8.858	7.430	433.5E6	349.5E6	437.589m	445.529
34) L7 AR-1260-4	9.102	7.913	429.4E6	306.9E6	446.354m	474.786
35) L7 AR-1260-5	9.446	8.158	983.7E6	841.5E6	433.280	474.071

AJ
 06/02/21

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.