

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ052919\
Data File : PQ040374.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 May 2019 21:05
Operator : SM\AJ
Sample : K3078-20
Misc :
ALS Vial : 32 Sample Multiplier: 1


Instrument :
ECD_Q
ClientSampled :
ETMW7

Manual Integrations
APPROVED

Ankita
5/30/2019 11:53:16 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 30 03:50:58 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ052219CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu May 23 04:10:58 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(21) AR-1248-1 (L5)
R.T. Response Conc
7.07 19279655 249.08
7.35 65616249 582.57
7.62 24544106 177.82
8.07 62494249 413.95
8.07 62494249 415.08
(21) AR-1248-1 #2 (L5)
R.T. Response Conc
6.20 13644800 236.79
6.46 51037213 580.50
6.51 36271511 400.86
6.71 56208892 509.00
7.17 39750672 350.12

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ052919\
Data File : PQ040374.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 May 2019 21:05
Operator : SM\AJ
Sample : K3078-20
Misc :
ALS Vial : 32 Sample Multiplier: 1


Instrument :
ECD_Q
ClientSampleId :
ETMW7

Manual Integrations
APPROVED

Ankita
5/30/2019 11:53:16 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 30 03:50:58 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ052219CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu May 23 04:10:58 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(21) AR-1248-1 #2 (L5)
R.T. Response Conc
7.04 6767287 87.43
7.35 62216988 552.39
7.58 65630623 475.49
8.03 35575208 235.64
8.07 59590072 395.79

(21) AR-1248-1 #2 (L5)
R.T. Response Conc
6.17 4999321 86.76
6.46 51037213 580.50
6.51 36271511 400.86
6.71 56208892 509.00
7.17 39750672 350.12

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ052919\
Data File : PQ040374.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 May 2019 21:05
Operator : SM\AJ
Sample : K3078-20
Misc :
ALS Vial : 32 Sample Multiplier: 1


Instrument :
ECD_Q
ClientSampleId :
ETMW7

Manual Integrations
APPROVED

Ankita
5/30/2019 11:53:16 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 30 03:50:58 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ052219CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu May 23 04:10:58 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(26) AR-1254-1 (L6)
R.T. Response Conc
8.03 35923691 205.99
8.24 102167989 369.43
8.63 99653578 335.78
8.93 47797564 229.99
9.40 6984214 24.25
(26) AR-1254-1 #2 (L6)
R.T. Response Conc
7.12 61854484 322.94
7.29 46981173 270.43
7.74 97065928 335.89
7.99 47192287 264.31
8.43 47319154 183.67

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ052919\
Data File : PQ040374.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 May 2019 21:05
Operator : SM\AJ
Sample : K3078-20
Misc :
ALS Vial : 32 Sample Multiplier: 1


Instrument :
ECD_Q
ClientSampleId :
ETMW7

Manual Integrations
APPROVED

Ankita
5/30/2019 11:53:16 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 30 03:50:58 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ052219CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu May 23 04:10:58 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(26) AR-1254-1 #2 (L6)
R.T. Response Conc
7.99 58011689 332.65
8.24 104127497 376.52
8.63 99653578 335.78
8.93 47797564 229.99
9.37 49099750 170.49
(26) AR-1254-1 #2 (L6)
R.T. Response Conc
7.12 61854484 322.94
7.29 46981173 270.43
7.74 97065928 335.89
7.99 39189547 219.49
8.43 47319154 183.67

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ052919\
 Data File : PQ040374.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 May 2019 21:05
 Operator : SM\AJ
 Sample : K3078-20
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 ETMW7

Manual Integrations
 APPROVED

Ankita
 5/30/2019 11:53:16 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 30 03:50:58 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ052219CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu May 23 04:10:58 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	5.590	4.731	36593733	28267683	11.171	13.161
2) SA Decachlor...	11.892	10.448	129.0E6	77378136	18.054	18.609
Target Compounds						
21) L5 AR-1248-1	7.037	6.173	6767287	4999321	87.430m	86.759m
22) L5 AR-1248-2	7.350	6.463	62216988	51037213	552.391m	580.499
23) L5 AR-1248-3	7.580	6.510	65630623	36271511	475.491m	400.863
24) L5 AR-1248-4	8.026	6.715	35575208	56208892	235.645m	509.003 #
25) L5 AR-1248-5	8.067	7.170	59590072	39750672	395.793m	350.116
26) L6 AR-1254-1	7.994	7.123	58011689	61854484	332.647m	322.939
27) L6 AR-1254-2	8.235	7.291	104.1E6	46981173	376.519m	270.426 #
28) L6 AR-1254-3	8.628	7.738	99653578	97065928	335.778	335.889
29) L6 AR-1254-4	8.930	7.988	47797564	39189547	229.985	219.491m
30) L6 AR-1254-5	9.366	8.431	49099750	47319154	170.489m	183.671

13
 06/05/19

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.