

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ053023\
 Data File : PQ061328.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 May 2023 16:18
 Operator : YP\AJ
 Sample : 02955-12
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 WC21

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 05/31/2023
 Supervised By :Ankita Jodhani 05/31/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 30 21:05:31 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ052523.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu May 25 13:25:08 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.405	2.761	63235300	44970839	12.620	15.567
2) SA Decachlor...	8.583	7.534	27873255	32545651	7.206m	8.938

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ053023\
Data File : PQ061328.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 May 2023 16:18
Operator : YP\AJ
Sample : 02955-12
Misc :
ALS Vial : 20 Sample Multiplier: 1



Instrument :
ECD_Q
ClientSampleId :
WC21


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/31/2023
Supervised By :Ankita Jodhani 05/31/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 30 21:05:31 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ052523.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu May 25 13:25:08 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.405 min
Delta R.T.: 0.000 min
Response: 63235300
Conc: 12.62 ng/ml

Instrument :
ECD_Q
ClientSampleId :
WC21


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/31/2023
Supervised By :Ankita Jodhani 05/31/2023




#1 Tetrachloro-m-xylene

R.T.: 2.761 min
Delta R.T.: 0.000 min
Response: 44970839
Conc: 15.57 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.583 min
Delta R.T.: -0.002 min
Response: 27873255
Conc: 7.21 ng/ml m



#2 Decachlorobiphenyl

R.T.: 7.534 min
Delta R.T.: 0.000 min
Response: 32545651
Conc: 8.94 ng/ml