

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_Q\Data\PQ053122\  
 Data File : PQ057881.D  
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch  
 Acq On : 31 May 2022 14:49  
 Operator : YP\AJ  
 Sample : N2930-19DL 5X  
 Misc :  
 ALS Vial : 23 Sample Multiplier: 1

Instrument :  
 ECD\_Q  
 ClientSampleId :  
 P006-SS002-1218-01DL

Manual Integrations  
 APPROVED

Reviewed By :Yogesh Patel 06/01/2022  
 Supervised By :Ankita Jodhani 06/01/2022

Integration File signal 1: autoint1.e  
 Integration File signal 2: autoint2.e  
 Quant Time: Jun 01 03:42:06 2022  
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_Q\Method\PQ052822.M  
 Quant Title : GC EXTRACTABLES  
 QLast Update : Sat May 28 06:33:00 2022  
 Response via : Initial Calibration  
 Integrator: ChemStation

Volume Inj. : 2 µl  
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2  
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

| Compound                    | RT#1   | RT#2  | Resp#1   | Resp#2   | ng/ml    | ng/ml     |
|-----------------------------|--------|-------|----------|----------|----------|-----------|
| -----                       |        |       |          |          |          |           |
| System Monitoring Compounds |        |       |          |          |          |           |
| 1) SA Tetrachlo...          | 4.690  | 4.054 | 51063375 | 44285198 | 3.538    | 3.897     |
| 2) SA Decachlor...          | 10.615 | 9.209 | 46985438 | 33729010 | 3.746    | 3.554     |
| Target Compounds            |        |       |          |          |          |           |
| 31) L7 AR-1260-1            | 7.483  | 6.659 | 26711738 | 36432326 | 61.071   | 70.123    |
| 32) L7 AR-1260-2            | 7.739  | 6.844 | 106.8E6  | 101.4E6  | 192.743  | 162.534   |
| 33) L7 AR-1260-3            | 8.027  | 7.004 | 148.2E6  | 74054435 | 219.986m | 113.693 # |
| 34) L7 AR-1260-4            | 8.340  | 7.478 | 95767763 | 88806399 | 191.922  | 176.610   |
| 35) L7 AR-1260-5            | 8.681  | 7.715 | 227.5E6  | 238.7E6  | 204.160  | 194.357   |
| -----                       |        |       |          |          |          |           |

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_Q\Data\PQ053122\  
Data File : PQ057881.D  
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch  
Acq On : 31 May 2022 14:49  
Operator : YP\AJ  
Sample : N2930-19DL 5X  
Misc :  
ALS Vial : 23 Sample Multiplier: 1



Instrument :  
ECD\_Q  
ClientSampleId :  
P006-SS002-1218-01DL


Manual Integrations  
APPROVED

Reviewed By :Yogesh Patel 06/01/2022  
Supervised By :Ankita Jodhani 06/01/2022

Integration File signal 1: autoint1.e  
Integration File signal 2: autoint2.e  
Quant Time: Jun 01 03:42:06 2022  
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_Q\Method\PQ052822.M  
Quant Title : GC EXTRACTABLES  
QLast Update : Sat May 28 06:33:00 2022  
Response via : Initial Calibration  
Integrator: ChemStation

Volume Inj. : 2 µl  
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2  
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.690 min  
Delta R.T.: 0.000 min  
Response: 51063375  
Conc: 3.54 ng/ml

Instrument :  
ECD\_Q  
ClientSampleId :  
P006-SS002-1218-01DL


Manual Integrations  
APPROVED

Reviewed By :Yogesh Patel 06/01/2022  
Supervised By :Ankita Jodhani 06/01/2022




#1 Tetrachloro-m-xylene

R.T.: 4.054 min  
Delta R.T.: 0.001 min  
Response: 44285198  
Conc: 3.90 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.615 min  
Delta R.T.: 0.000 min  
Response: 46985438  
Conc: 3.75 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.209 min  
Delta R.T.: 0.000 min  
Response: 33729010  
Conc: 3.55 ng/ml




#31 AR-1260-1

R.T.: 7.483 min  
Delta R.T.: 0.001 min  
Response: 26711738  
Conc: 61.07 ng/ml

Instrument :  
ECD\_Q  
ClientSampleId :  
P006-SS002-1218-01DL


Manual Integrations  
APPROVED

Reviewed By :Yogesh Patel 06/01/2022  
Supervised By :Ankita Jodhani 06/01/2022




#31 AR-1260-1

R.T.: 6.659 min  
Delta R.T.: 0.000 min  
Response: 36432326  
Conc: 70.12 ng/ml




#32 AR-1260-2

R.T.: 7.739 min  
Delta R.T.: 0.000 min  
Response: 106799308  
Conc: 192.74 ng/ml



#32 AR-1260-2

R.T.: 6.844 min  
Delta R.T.: 0.000 min  
Response: 101360596  
Conc: 162.53 ng/ml




#33 AR-1260-3

R.T.: 8.027 min  
Delta R.T.: 0.000 min  
Response: 148179633  
Conc: 219.99 ng/ml

Instrument :  
ECD\_Q  
ClientSampleId :  
P006-SS002-1218-01DL


Manual Integrations  
APPROVED

Reviewed By :Yogesh Patel 06/01/2022  
Supervised By :Ankita Jodhani 06/01/2022




#33 AR-1260-3

R.T.: 7.004 min  
Delta R.T.: 0.000 min  
Response: 74054435  
Conc: 113.69 ng/ml




#34 AR-1260-4

R.T.: 8.340 min  
Delta R.T.: 0.001 min  
Response: 95767763  
Conc: 191.92 ng/ml



#34 AR-1260-4

R.T.: 7.478 min  
Delta R.T.: 0.000 min  
Response: 88806399  
Conc: 176.61 ng/ml




#35 AR-1260-5

R.T.: 8.681 min  
Delta R.T.: 0.001 min  
Response: 227459601  
Conc: 204.16 ng/ml

Instrument :  
ECD\_Q  
ClientSampleId :  
P006-SS002-1218-01DL

Manual Integrations  
APPROVED

Reviewed By :Yogesh Patel 06/01/2022  
Supervised By :Ankita Jodhani 06/01/2022



#35 AR-1260-5

R.T.: 7.715 min  
Delta R.T.: 0.000 min  
Response: 238670399  
Conc: 194.36 ng/ml