

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_Q\Data\PQ053122\  
 Data File : PQ057883.D  
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch  
 Acq On : 31 May 2022 15:46  
 Operator : YP\AJ  
 Sample : I.BLK  
 Misc :  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 ECD\_Q  
 ClientSampleId :  
 I.BLK

Integration File signal 1: autoint1.e  
 Integration File signal 2: autoint2.e  
 Quant Time: Jun 01 03:43:26 2022  
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_Q\Method\PQ052822.M  
 Quant Title : GC EXTRACTABLES  
 QLast Update : Sat May 28 06:33:00 2022  
 Response via : Initial Calibration  
 Integrator: ChemStation

Volume Inj. : 2 µl  
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2  
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

|                             | Compound        | RT#1   | RT#2  | Resp#1  | Resp#2  | ng/ml  | ng/ml  |
|-----------------------------|-----------------|--------|-------|---------|---------|--------|--------|
| -----                       |                 |        |       |         |         |        |        |
| System Monitoring Compounds |                 |        |       |         |         |        |        |
| 1)                          | SA Tetrachlo... | 4.690  | 4.053 | 298.1E6 | 275.3E6 | 20.654 | 24.224 |
| 2)                          | SA Decachlor... | 10.615 | 9.208 | 283.9E6 | 201.1E6 | 22.640 | 21.193 |

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_Q\Data\PQ053122\  
Data File : PQ057883.D  
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch  
Acq On : 31 May 2022 15:46  
Operator : YP\AJ  
Sample : I.BLK  
Misc :  
ALS Vial : 2 Sample Multiplier: 1

Instrument :  
ECD\_Q  
ClientSampleId :  
I.BLK

Integration File signal 1: autoint1.e  
Integration File signal 2: autoint2.e  
Quant Time: Jun 01 03:43:26 2022  
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_Q\Method\PQ052822.M  
Quant Title : GC EXTRACTABLES  
QLast Update : Sat May 28 06:33:00 2022  
Response via : Initial Calibration  
Integrator: ChemStation

Volume Inj. : 2 µl  
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2  
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.690 min  
Delta R.T.: 0.000 min  
Response: 298077977  
Conc: 20.65 ng/ml

Instrument :  
ECD\_Q  
ClientSampleId :  
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 4.053 min  
Delta R.T.: 0.000 min  
Response: 275255255  
Conc: 24.22 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.615 min  
Delta R.T.: 0.000 min  
Response: 283949545  
Conc: 22.64 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.208 min  
Delta R.T.: -0.002 min  
Response: 201115810  
Conc: 21.19 ng/ml