

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ060121\
 Data File : PQ053751.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Jun 2021 20:12
 Operator : DD\AJ
 Sample : AR1232ICC1600
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR12325001

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 02 07:33:14 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ060121CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jun 02 07:30:58 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	5.217	4.233	1814.8E6	997.4E6	72.606	77.270
2) SA Decachlor...	11.401	9.567	3624.9E6	2357.3E6	144.219	144.707
Target Compounds						
11) L3 AR-1232-1	5.655	4.678	672.6E6	400.3E6	1433.993	1463.988
12) L3 AR-1232-2	6.247	5.511	375.5E6	450.3E6	1436.409	1463.762
13) L3 AR-1232-3	6.566	5.703	803.3E6	225.4E6	1412.739	1419.238
14) L3 AR-1232-4	6.741	5.798	390.3E6	191.8E6	1425.879	1402.653
15) L3 AR-1232-5	6.834	5.983	268.8E6	213.1E6	1407.332	1447.763

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ060121\
Data File : PQ053751.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Jun 2021 20:12
Operator : DD\AJ
Sample : AR1232ICC1600
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR12325001

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 02 07:33:14 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ060121CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jun 02 07:30:58 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

