

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ060121\
 Data File : PQ053775.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Jun 2021 02:45
 Operator : DD\AJ
 Sample : AR1268ICC800
 Misc :
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR12684001

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 02 08:48:26 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ060121CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jun 02 08:46:41 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	5.218	4.233	1155.3E6	632.7E6	39.537	38.331
2) SA Decachlor...	11.400	9.565	4448.0E6	2837.0E6	77.831	75.982
Target Compounds						
41) L9 AR-1268-1	9.777	8.444	4416.2E6	3288.1E6	787.232	758.418
42) L9 AR-1268-2	9.879	8.511	3938.6E6	2779.0E6	785.696	759.521
43) L9 AR-1268-3	10.122	8.718	3524.8E6	2433.5E6	787.680	763.136
44) L9 AR-1268-4	10.591	9.011	1386.7E6	965.2E6	785.075	772.207
45) L9 AR-1268-5	11.038	9.308	10943.8E6	6937.7E6	790.950	753.393

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ060121\
Data File : PQ053775.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Jun 2021 02:45
Operator : DD\AJ
Sample : AR1268ICC800
Misc :
ALS Vial : 41 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampled :
AR12684001

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 02 08:48:26 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ060121CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jun 02 08:46:41 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

