

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ060324\
 Data File : PQ066808.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 03 Jun 2024 09:23
 Operator : YP\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR1660CCC500

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 06/04/2024
 Supervised By :Ankita Jodhani 06/04/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 03 11:22:00 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ053124.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Jun 03 10:00:54 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.453	2.826	281.2E6	302.1E6	50.778	51.647m
2) SA Decachlor...	8.727	7.617	188.1E6	271.5E6	49.600	47.269
Target Compounds						
3) L1 AR-1016-1	4.566	3.835	85211572	102.0E6	503.390	498.758
4) L1 AR-1016-2	4.585	3.852	135.2E6	148.0E6	513.728	498.727
5) L1 AR-1016-3	4.643	4.013	82927189	82684775	516.702	503.727
6) L1 AR-1016-4	4.735	4.062	65767440	68734047	513.198	499.013
7) L1 AR-1016-5	5.027	4.258	68288090	85900331	530.220	500.610
31) L7 AR-1260-1	6.147	5.258	132.4E6	163.4E6	513.265	493.755
32) L7 AR-1260-2	6.412	5.447	160.6E6	192.1E6	528.176	491.484
33) L7 AR-1260-3	6.773	5.590	114.8E6	184.6E6	518.570	487.242
34) L7 AR-1260-4	6.994	6.052	124.7E6	153.4E6	514.727	488.122
35) L7 AR-1260-5	7.312	6.298	251.6E6	320.4E6	514.166	485.431

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ060324\
Data File : PQ066808.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Jun 2024 09:23
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1660CCC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/04/2024
Supervised By :Ankita Jodhani 06/04/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 03 11:22:00 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ053124.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Jun 03 10:00:54 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

