

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ060422\
 Data File : PQ057979.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 03 Jun 2022 14:21
 Operator : YP\AJ
 Sample : N2988-22
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 P015-SS005-1824-01

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 06/06/2022
 Supervised By :Ankita Jodhani 06/06/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 03 23:12:39 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ052822.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat May 28 06:33:00 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.688	4.051	257.2E6	251.3E6	17.825	22.114
2) SA Decachlor...	10.611	9.201	183.4E6	145.8E6	14.620	15.359
Target Compounds						
31) L7 AR-1260-1	7.480	6.655	8745817	16637707	19.996	32.023 #
32) L7 AR-1260-2	7.738	6.840	21043680	28520133	37.978	45.733
33) L7 AR-1260-3	8.024	7.000	26959284	19934098	40.024m	30.604
34) L7 AR-1260-4	8.336	7.473	21639312	21145433	43.366	42.052
35) L7 AR-1260-5	8.677	7.710	67399149	73667081	60.495	59.989

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ060422\
Data File : PQ057979.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Jun 2022 14:21
Operator : YP\AJ
Sample : N2988-22
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
P015-SS005-1824-01

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/06/2022
Supervised By :Ankita Jodhani 06/06/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 03 23:12:39 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ052822.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat May 28 06:33:00 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

