

Method Path : Z:\PESTPCBSRV\HPCHEM1\ECD_Q\Method\
 Method File : PQ060618.M
 Title : GC EXTRACTABLES
 Last Update : Thu Jun 07 01:13:06 2018
 Response Via : Initial Calibration

Calibration Files

500 =PQ029621.D 1000 =PQ029619.D 750 =PQ029620.D
 250 =PQ029622.D 50 =PQ029623.D

Compound		500	1000	750	250	50	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	1.392	1.791	1.579	1.473	1.874	1.622	E8	12.68
2) SA	Decachlorobiphenyl	1.245	1.329	1.238	1.223	1.266	1.260	E8	3.31
3) L1	AR-1016-1	1.984	2.372	2.210	2.099	2.447	2.222	E6	8.57
4) L1	AR-1016-2	2.438	2.997	2.753	2.522	3.043	2.751	E6	9.89
5) L1	AR-1016-3	3.262	3.942	3.613	3.429	3.688	3.587	E6	7.21
6) L1	AR-1016-4	2.108	2.556	2.355	2.128	2.766	2.383	E6	11.84
7) L1	AR-1016-5	2.721	3.186	2.980	2.909	3.911	3.141	E6	14.69
8) L2	AR-1221-1	2.245					2.245	E6	0.00
9) L2	AR-1221-2	1.591					1.591	E6	0.00
10) L2	AR-1221-3	4.995					4.995	E6	0.00
11) L3	AR-1232-1	2.379					2.379	E6	0.00
12) L3	AR-1232-2	1.227					1.227	E6	0.00
13) L3	AR-1232-3	1.662					1.662	E6	0.00
14) L3	AR-1232-4	1.338					1.338	E6	0.00
15) L3	AR-1232-5	1.171					1.171	E6	0.00
16) L4	AR-1242-1	2.329					2.329	E6	0.00
17) L4	AR-1242-2	3.357					3.357	E6	0.00
18) L4	AR-1242-3	3.063					3.063	E6	0.00
19) L4	AR-1242-4	2.411					2.411	E6	0.00
20) L4	AR-1242-5	2.605					2.605	E6	0.00
21) L5	AR-1248-1	3.820					3.820	E6	0.00
22) L5	AR-1248-2	3.260					3.260	E6	0.00
23) L5	AR-1248-3	4.513					4.513	E6	0.00
24) L5	AR-1248-4	4.530					4.530	E6	0.00
25) L5	AR-1248-5	4.314					4.314	E6	0.00
26) L6	AR-1254-1	4.380					4.380	E6	0.00
27) L6	AR-1254-2	6.817					6.817	E6	0.00
28) L6	AR-1254-3	7.335					7.335	E6	0.00
29) L6	AR-1254-4	5.524					5.524	E6	0.00
30) L6	AR-1254-5	4.050					4.050	E6	0.00
31) L7	AR-1260-1	4.535	5.420	5.010	4.743	5.655	5.073	E6	9.15
32) L7	AR-1260-2	5.902	6.935	6.498	5.978	7.251	6.513	E6	9.03
33) L7	AR-1260-3	7.864	8.212	7.696	7.986	8.365	8.025	E6	3.33
34) L7	AR-1260-4	1.234	1.397	1.313	1.229	1.347	1.304	E7	5.57
35) L7	AR-1260-5	6.960	7.705	7.321	6.926	7.469	7.276	E6	4.59
36) L8	AR-1262-1	4.390					4.390	E6	0.00
37) L8	AR-1262-2	5.474					5.474	E6	0.00
38) L8	AR-1262-3	8.400					8.400	E6	0.00
39) L8	AR-1262-4	6.990					6.990	E6	0.00
40) L8	AR-1262-5	1.643					1.643	E7	0.00
41) L9	AR-1268-1	1.728					1.728	E7	0.00
42) L9	AR-1268-2	1.618					1.618	E7	0.00
43) L9	AR-1268-3	1.436					1.436	E7	0.00
44) L9	AR-1268-4	5.548					5.548	E6	0.00
45) L9	AR-1268-5	4.522					4.522	E7	0.00

Signal #2 Calibration Files

500 =PQ029621.D 1000 =PQ029619.D 750 =PQ029620.D
 250 =PQ029622.D 50 =PQ029623.D

Compound		500	1000	750	250	50	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	0.793	0.957	0.877	0.805	1.080	0.902	E7	13.18
2) SA	Decachlorobiphenyl	1.219	1.200	1.210	1.282	1.580	1.298	E7	12.38
3) L1	AR-1016-1	1.985	2.299	2.154	2.074	2.615	2.225	E5	11.07
4) L1	AR-1016-2	2.059	2.427	2.267	2.188	2.591	2.306	E5	9.00
5) L1	AR-1016-3	1.644	1.884	1.785	1.779	2.521	1.923	E5	17.96

Method Path : Z:\PESTPCBSRV\HPCHEM1\ECD_Q\Method\
 Method File : PQ060618.M
 Title : GC EXTRACTABLES
 Last Update : Thu Jun 07 01:13:06 2018
 Response Via : Initial Calibration

Calibration Files

500 =PQ029621.D 1000 =PQ029619.D 750 =PQ029620.D
 250 =PQ029622.D 50 =PQ029623.D

	Compound	500	1000	750	250	50	Avg		%RSD
6)	L1 AR-1016-4	2.084	2.383	2.255	2.187	3.222	2.426	E5	18.87
7)	L1 AR-1016-5	2.242	2.562	2.429	2.339	3.344	2.583	E5	17.08
8)	L2 AR-1221-1	1.473					1.473	E5	0.00
9)	L2 AR-1221-2	1.095					1.095	E5	0.00
10)	L2 AR-1221-3	3.382					3.382	E5	0.00
11)	L3 AR-1232-1	1.674					1.674	E5	0.00
12)	L3 AR-1232-2	1.539					1.539	E5	0.00
13)	L3 AR-1232-3	1.038					1.038	E5	0.00
14)	L3 AR-1232-4	1.087					1.087	E5	0.00
15)	L3 AR-1232-5	1.283					1.283	E5	0.00
16)	L4 AR-1242-1	1.901					1.901	E5	0.00
17)	L4 AR-1242-2	4.098					4.098	E5	0.00
18)	L4 AR-1242-3	1.986					1.986	E5	0.00
19)	L4 AR-1242-4	2.035					2.035	E5	0.00
20)	L4 AR-1242-5	2.196					2.196	E5	0.00
21)	L5 AR-1248-1	2.657					2.657	E5	0.00
22)	L5 AR-1248-2	2.808					2.808	E5	0.00
23)	L5 AR-1248-3	3.375					3.375	E5	0.00
24)	L5 AR-1248-4	5.098					5.098	E5	0.00
25)	L5 AR-1248-5	3.698					3.698	E5	0.00
26)	L6 AR-1254-1	5.546					5.546	E5	0.00
27)	L6 AR-1254-2	4.830					4.830	E5	0.00
28)	L6 AR-1254-3	8.517					8.517	E5	0.00
29)	L6 AR-1254-4	5.622					5.622	E5	0.00
30)	L6 AR-1254-5	3.502					3.502	E5	0.00
31)	L7 AR-1260-1	4.970	5.537	5.330	5.320	7.564	5.744	E5	18.06
32)	L7 AR-1260-2	5.919	6.556	6.319	6.236	8.351	6.676	E5	14.44
33)	L7 AR-1260-3	5.197	5.610	5.454	5.494	7.107	5.773	E5	13.18
34)	L7 AR-1260-4	1.370	1.473	1.419	1.424	1.778	1.493	E6	10.95
35)	L7 AR-1260-5	1.009	1.063	1.039	1.057	1.293	1.092	E6	10.44
36)	L8 AR-1262-1	4.942					4.942	E5	0.00
37)	L8 AR-1262-2	6.098					6.098	E5	0.00
38)	L8 AR-1262-3	9.082					9.082	E5	0.00
39)	L8 AR-1262-4	8.080					8.080	E5	0.00
40)	L8 AR-1262-5	1.735					1.735	E6	0.00
41)	L9 AR-1268-1	1.762					1.762	E6	0.00
42)	L9 AR-1268-2	1.659					1.659	E6	0.00
43)	L9 AR-1268-3	1.398					1.398	E6	0.00
44)	L9 AR-1268-4	6.311					6.311	E5	0.00
45)	L9 AR-1268-5	4.230					4.230	E6	0.00

(#) = Out of Range