

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ060618\
 Data File : PQ029625.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Jun 2018 15:41
 Operator : UA/SM
 Sample : AR1232ICC500
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR1232ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 07 00:56:19 2018
 Quant Method : Z:\PESTPCBSRV\HPCHEM1\ECD_Q\Method\PQ060618.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jun 07 00:55:48 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.593	3.762	7482.8E6	428.0E6	45.556	46.779
2) SA Decachlor...	10.316	8.696	5549.7E6	560.1E6	45.003	43.814
Target Compounds						
11) L3 AR-1232-1	4.955	4.133	1189.3E6	83723248	500.000	500.000
12) L3 AR-1232-2	5.479	4.906	613.3E6	76953968	500.000	500.000
13) L3 AR-1232-3	5.826	5.023	831.0E6	51896133	500.000	500.000
14) L3 AR-1232-4	5.923	5.273	669.0E6	54352313	500.000	500.000
15) L3 AR-1232-5	6.213	5.618	585.3E6	64156355	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ060618\
Data File : PQ029625.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Jun 2018 15:41
Operator : UA/SM
Sample : AR1232ICC500
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1232ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 07 00:56:19 2018
Quant Method : Z:\PESTPCBSRV\HPCHEM1\ECD_Q\Method\PQ060618.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jun 07 00:55:48 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm

