

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ060619\
 Data File : PQ040648.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Jun 2019 19:00
 Operator : SM\AJ
 Sample : AR1254ICC400
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR1254301

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 06 19:16:06 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ060619CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jun 06 19:13:45 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	5.541	4.645	48475593	37245611	20.000	20.000
2)	SA Decachlor...	11.815	10.337	296.8E6	181.0E6	40.000	40.000
Target Compounds							
26)	L6 AR-1254-1	7.946	7.035	54514428	68047531	400.000	400.000
27)	L6 AR-1254-2	8.182	7.203	92459730	60845371	400.000	400.000
28)	L6 AR-1254-3	8.578	7.649	101.9E6	101.7E6	400.000	400.000
29)	L6 AR-1254-4	8.882	7.900	82262268	61937336	400.000	400.000
30)	L6 AR-1254-5	9.320	8.342	95872413	98813987	400.000	400.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ060619\
Data File : PQ040648.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Jun 2019 19:00
Operator : SM\AJ
Sample : AR1254ICC400
Misc :
ALS Vial : 30 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampled :
AR1254301

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 06 19:16:06 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ060619CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jun 06 19:13:45 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

