Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ060619\
Data File : PQ040647.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Jun 2019 18:42
Operator : SM\AJ
Sample : AR1254ICC200
Misc :
ALS Vial : 29 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1254201

Manual Integrations APPROVED
Ankita
6/10/2019 12:26:59 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 06 19:19:11 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ060619CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jun 06 19:13:45 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	5.541	4.646	25248281	19272265	10.204	10.171
2)	SA Decachlor...	11.815	10.338	153.8E6	92075387	20.362	20.175
	Target Compounds						
26)	L6 AR-1254-1	7.946	7.035	28396348	33754478	204.094	199.205
27)	L6 AR-1254-2	8.183	7.204	45396321	30381242	198.181	199.864
28)	L6 AR-1254-3	8.579	7.649	49608061	49136484	197.369	196.571
29)	L6 AR-1254-4	8.882	7.900	41568690	36101008	201.058	215.305
30)	L6 AR-1254-5	9.320	8.342	46742728	47455605	196.115m	195.971

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ060619\
Data File : PQ040647.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Jun 2019 18:42
Operator : SM\AJ
Sample : AR1254ICC200
Misc :
ALS Vial : 29 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1254201

Manual Integrations APPROVED
Ankita
6/10/2019 12:26:59 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 06 19:19:11 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ060619CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jun 06 19:13:45 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm