

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ060622\
 Data File : PQ057985.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Jun 2022 12:45
 Operator : YP\AJ
 Sample : N2930-12DL 2X
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 P005-SS004-0006-01DL

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 07 02:56:54 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ052822.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat May 28 06:33:00 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.691	4.052	144.1E6	136.0E6	9.987	11.965
2) SA Decachlor...	10.614	9.202	158.9E6	146.0E6	12.667	15.385
Target Compounds						
31) L7 AR-1260-1	7.483	6.656	174.3E6	232.9E6	398.575	448.262
32) L7 AR-1260-2	7.739	6.841	533.7E6	544.3E6	963.247	872.725
33) L7 AR-1260-3	8.026	6.999	788.7E6	334.2E6	1170.950	513.025 #
34) L7 AR-1260-4	8.338	7.474	534.3E6	458.4E6	1070.752	911.587
35) L7 AR-1260-5	8.680	7.711	1525.9E6	1327.2E6	1369.621	1080.796

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ060622\
Data File : PQ057985.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Jun 2022 12:45
Operator : YP\AJ
Sample : N2930-12DL 2X
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
P005-SS004-0006-01DL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 07 02:56:54 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ052822.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat May 28 06:33:00 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

