

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ060624\
 Data File : PQ066977.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Jun 2024 15:28
 Operator : YP\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR1660CCC500

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 06/07/2024
 Supervised By :Ankita Jodhani 06/07/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 06 15:40:24 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ053124.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Jun 03 10:00:54 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.448	2.825	306.9E6	307.9E6	55.414	52.629
2)	SA Decachlor...	8.703	7.611	206.4E6	273.3E6	54.425	47.592m
Target Compounds							
3)	L1 AR-1016-1	4.558	3.833	93303242	105.8E6	551.192	517.620
4)	L1 AR-1016-2	4.577	3.850	142.5E6	153.0E6	541.541	515.753
5)	L1 AR-1016-3	4.635	4.011	86827847	84663691	541.006	515.783
6)	L1 AR-1016-4	4.728	4.059	71136475	70566726	555.093	512.318
7)	L1 AR-1016-5	5.018	4.256	69204488	87323029	537.336	508.901
31)	L7 AR-1260-1	6.136	5.255	141.3E6	165.6E6	547.820	500.373
32)	L7 AR-1260-2	6.400	5.443	170.5E6	197.4E6	560.782	505.084m
33)	L7 AR-1260-3	6.759	5.587	123.1E6	187.7E6	556.438	495.370
34)	L7 AR-1260-4	6.979	6.050	136.2E6	156.4E6	562.058	497.864
35)	L7 AR-1260-5	7.297	6.295	274.2E6	333.7E6	560.402	505.488

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ060624\
Data File : PQ066977.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Jun 2024 15:28
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1660CCC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/07/2024
Supervised By :Ankita Jodhani 06/07/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 06 15:40:24 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ053124.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Jun 03 10:00:54 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

