Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ060819\
Data File : PQ040741.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Jun 2019 08:44
Operator : SM\AJ
Sample : AR1248CCC400
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1248304

Manual Integrations APPROVED
Ankita
6/10/2019 1:30:11 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 08 21:47:09 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ060619CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jun 07 01:39:59 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	5.525	4.634	52459108	36892759	22.210	22.090
2)	SA Decachlor...	11.784	10.323	298.9E6	156.2E6	39.687	35.708
	Target Compounds						
21)	L5 AR-1248-1	6.972	6.076	21564343	17075350	460.704m	435.793m
22)	L5 AR-1248-2	7.284	6.365	31247301	25151727	434.526	412.447
23)	L5 AR-1248-3	7.513	6.413	37659450	25117158	434.764m	411.118
24)	L5 AR-1248-4	7.960	6.616	47339778	31857103	454.660m	409.819
25)	L5 AR-1248-5	8.001	7.070	43789377	32268277	451.396	405.969m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ060819\
Data File : PQ040741.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Jun 2019 08:44
Operator : SM\AJ
Sample : AR1248CCC400
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1248304

Manual Integrations APPROVED
Ankita
6/10/2019 1:30:11 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 08 21:47:09 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ060619CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jun 07 01:39:59 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm