

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ060821\
 Data File : PQ053834.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 08 Jun 2021 21:10
 Operator : DD\AJ
 Sample : AR1242CCC400
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR12423018

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 09 01:55:32 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ060121CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jun 02 09:01:47 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	5.218	4.236	510.4E6	283.4E6	21.062	22.352
2)	SA Decachlor...	11.399	9.566	604.3E6	341.1E6	24.654	22.422
Target Compounds							
16)	L4 AR-1242-1	6.542	5.491	244.3E6	194.1E6	320.294	457.107 #
17)	L4 AR-1242-2	6.566	5.511	368.9E6	280.8E6	306.549	435.233 #
18)	L4 AR-1242-3	6.632	5.703	214.8E6	144.0E6	294.326	439.001 #
19)	L4 AR-1242-4	6.741	5.799	172.6E6	134.9E6	295.108	427.888 #
20)	L4 AR-1242-5	7.523	6.362	176.8E6	163.3E6	269.870	378.705 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ060821\
Data File : PQ053834.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Jun 2021 21:10
Operator : DD\AJ
Sample : AR1242CCC400
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampled :
AR12423018

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 09 01:55:32 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ060121CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jun 02 09:01:47 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

