

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ060923\
 Data File : PQ061341.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Jun 2023 13:21
 Operator : YP\AJ
 Sample : AR1242ICC750
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR1242ICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 09 15:00:37 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ060923.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jun 09 14:59:43 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.405	2.761	334.5E6	198.7E6	73.361	73.618
2) SA Decachlor...	8.585	7.535	246.3E6	248.4E6	72.650	73.216
Target Compounds						
16) L4 AR-1242-1	4.506	3.764	83877569	55416600	752.777	714.111
17) L4 AR-1242-2	4.525	3.780	124.7E6	82411237	726.844	726.396
18) L4 AR-1242-3	4.583	3.940	78487647	44168588	721.452	726.363
19) L4 AR-1242-4	4.674	4.025	63729255	43578162	719.215	721.005
20) L4 AR-1242-5	5.390	4.521	65014105	58200331	726.985	713.542



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ060923\
Data File : PQ061341.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Jun 2023 13:21
Operator : YP\AJ
Sample : AR1242ICC750
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1242ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 09 15:00:37 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ060923.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jun 09 14:59:43 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.405 min
Delta R.T.: 0.000 min
Response: 334534964
Conc: 73.36 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1242ICC750




#1 Tetrachloro-m-xylene

R.T.: 2.761 min
Delta R.T.: 0.000 min
Response: 198676373
Conc: 73.62 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.585 min
Delta R.T.: 0.000 min
Response: 246349264
Conc: 72.65 ng/ml



#2 Decachlorobiphenyl


R.T.: 7.535 min
Delta R.T.: 0.000 min
Response: 248414137
Conc: 73.22 ng/ml



#16 AR-1242-1


R.T.: 4.506 min
Delta R.T.: 0.000 min
Response: 83877569
Conc: 752.78 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1242ICC750




#16 AR-1242-1

R.T.: 3.764 min
Delta R.T.: 0.000 min
Response: 55416600
Conc: 714.11 ng/ml




#17 AR-1242-2

R.T.: 4.525 min
Delta R.T.: 0.000 min
Response: 124711949
Conc: 726.84 ng/ml



#17 AR-1242-2


R.T.: 3.780 min
Delta R.T.: 0.000 min
Response: 82411237
Conc: 726.40 ng/ml



#18 AR-1242-3


R.T.: 4.583 min
Delta R.T.: 0.000 min
Response: 78487647
Conc: 721.45 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1242ICC750




#18 AR-1242-3

R.T.: 3.940 min
Delta R.T.: 0.000 min
Response: 44168588
Conc: 726.36 ng/ml




#19 AR-1242-4

R.T.: 4.674 min
Delta R.T.: 0.000 min
Response: 63729255
Conc: 719.21 ng/ml



#19 AR-1242-4


R.T.: 4.025 min
Delta R.T.: 0.000 min
Response: 43578162
Conc: 721.01 ng/ml



#20 AR-1242-5

R.T.: 5.390 min
Delta R.T.: 0.000 min
Response: 65014105
Conc: 726.99 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1242ICC750



#20 AR-1242-5

R.T.: 4.521 min
Delta R.T.: 0.000 min
Response: 58200331
Conc: 713.54 ng/ml