

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ061219\
 Data File : PQ040802.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Jun 2019 16:58
 Operator : SM\AJ
 Sample : AR1242ICC1600
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR1242501

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 13 03:51:26 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ061219CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jun 13 03:49:57 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	5.740	4.798	280.5E6	262.0E6	75.416	79.005
2)	SA Decachlor...	12.176	10.588	1132.7E6	381.4E6	147.744	140.266
Target Compounds							
16)	L4 AR-1242-1	7.185	6.256	168.2E6	191.2E6	1459.020	1402.988
17)	L4 AR-1242-2	7.210	6.279	268.5E6	288.5E6	1445.863	1412.834
18)	L4 AR-1242-3	7.282	6.493	171.2E6	156.9E6	1535.802	1400.929
19)	L4 AR-1242-4	7.395	6.597	139.0E6	158.9E6	1548.742	1378.451
20)	L4 AR-1242-5	8.217	7.209	165.3E6	233.2E6	1488.042	1542.354

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ061219\
Data File : PQ040802.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Jun 2019 16:58
Operator : SM\AJ
Sample : AR1242ICC1600
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1242501

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 13 03:51:26 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ061219CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jun 13 03:49:57 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

