

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ061319\
Data File : PQ040835.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Jun 2019 09:42
Operator : SM\AJ
Sample : PB120555BS
Misc :
ALS Vial : 8 Sample Multiplier: 1



Instrument :
ECD_Q
ClientSampleId :
ALCS55

Manual Integrations
APPROVED

Ankita
6/20/2019 9:57:09 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 14 00:28:19 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ061219CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jun 13 05:59:16 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



QEdit

(3) AR-1016-1 (L1)

R.T.	Response	Conc
7.19	13694091	83.35
7.21	22020362	87.35
7.28	13236119	88.21
7.39	10043502	80.01
7.73	8398950	66.80

(3) AR-1016-1 #2 (L1)

R.T.	Response	Conc
6.26	16923966	87.83
6.28	25237254	90.23
6.49	13715874	89.92
6.55	12438666	99.28
6.80	16274527	96.31

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ061319\
Data File : PQ040835.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Jun 2019 09:42
Operator : SM\AJ
Sample : PB120555BS
Misc :
ALS Vial : 8 Sample Multiplier: 1



Instrument :
ECD_Q
ClientSampleId :
ALCS55

Manual Integrations
APPROVED

Ankita
6/20/2019 9:57:09 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 14 00:28:19 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ061219CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jun 13 05:59:16 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



QEdit

(3) AR-1016-1 #2 (L1)
R.T. Response Conc
7.19 13694091 83.35
7.21 22020362 87.35
7.28 13236119 88.21
7.39 10043502 80.01
7.73 10626600 84.52

(3) AR-1016-1 #2 (L1)
R.T. Response Conc
6.26 16923966 87.83
6.28 25237254 90.23
6.49 13715874 89.92
6.55 12438666 99.28
6.80 16274527 96.31

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ061319\
Data File : PQ040835.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Jun 2019 09:42
Operator : SM\AJ
Sample : PB120555BS
Misc :
ALS Vial : 8 Sample Multiplier: 1


Instrument :
ECD_Q
ClientSampleId :
ALCS55

Manual Integrations
APPROVED

Ankita
6/20/2019 9:57:09 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 14 00:28:19 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ061219CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jun 13 05:59:16 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(31) AR-1260-1 (L7)

R.T.	Response	Conc
8.95	27123991	105.34
9.22	34638544	103.05
9.59	26192023	90.86
9.84	28194698	93.44
10.19	57578339	84.17

(31) AR-1260-1 #2 (L7)

R.T.	Response	Conc
7.97	41914708	114.53
8.17	52685427	110.78
8.33	46453152	112.83
8.83	31985417	101.10
9.08	72625744	98.52

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ061319\
Data File : PQ040835.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Jun 2019 09:42
Operator : SM\AJ
Sample : PB120555BS
Misc :
ALS Vial : 8 Sample Multiplier: 1


Instrument :
ECD_Q
ClientSampleId :
ALCS55

Manual Integrations
APPROVED

Ankita
6/20/2019 9:57:09 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 14 00:28:19 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ061219CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jun 13 05:59:16 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(31) AR-1260-1 #2 (L7)
R.T. Response Conc
8.95 27123991 105.34
9.22 34638544 103.05
9.59 26192023 90.86
9.84 28194698 93.44
10.19 60823708 88.91

(31) AR-1260-1 #2 (L7)
R.T. Response Conc
7.97 41914708 114.53
8.17 52685427 110.78
8.33 46453152 112.83
8.83 31985417 101.10
9.08 72625744 98.52