

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ061622\
 Data File : PQ058134.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 16 Jun 2022 16:30
 Operator : YP\AJ
 Sample : AR1660CCC400
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR16603064

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 06/17/2022
 Supervised By :Ankita Jodhani 06/17/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 16 23:46:49 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ060822CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jun 09 05:05:02 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.687	4.049	261.1E6	266.7E6	17.619	17.307
2) SA Decachlor...	10.608	9.193	452.3E6	393.4E6	32.369	35.101
Target Compounds						
3) L1 AR-1016-1	5.872	5.154	136.0E6	168.1E6	344.122	339.562
4) L1 AR-1016-2	5.894	5.174	230.6E6	242.5E6	343.698	342.877
5) L1 AR-1016-3	5.957	5.354	144.3E6	122.7E6	349.263	333.116
6) L1 AR-1016-4	6.058	5.393	121.0E6	102.4E6	343.384	347.547
7) L1 AR-1016-5	6.352	5.611	107.3E6	131.9E6	349.911	345.823
31) L7 AR-1260-1	7.478	6.651	162.1E6	236.3E6	341.365	352.718
32) L7 AR-1260-2	7.735	6.835	187.5E6	275.9E6	325.796	342.846
33) L7 AR-1260-3	8.022	6.994	238.6E6	260.2E6	324.084	344.204m
34) L7 AR-1260-4	8.333	7.468	181.8E6	216.5E6	332.535	342.310
35) L7 AR-1260-5	8.674	7.705	371.0E6	492.8E6	311.863	333.457

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ061622\
Data File : PQ058134.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 16 Jun 2022 16:30
Operator : YP\AJ
Sample : AR1660CCC400
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR16603064

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/17/2022
Supervised By :Ankita Jodhani 06/17/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 16 23:46:49 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ060822CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jun 09 05:05:02 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

