

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ061721\
 Data File : PQ053916.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Jun 2021 12:15
 Operator : DD\AJ
 Sample : M2679-06MS
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 18 00:12:03 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ061621CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jun 17 06:07:32 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Instrument :
 ECD_Q
 ClientSampleId :
 C0CN8MS

Manual Integrations
 APPROVED

Ankita
 6/21/2021 10:33:22 AM

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	5.219	4.232	447.5E6	246.1E6	16.615	16.493
2)	SA Decachlor...	11.406	9.569	913.5E6	605.8E6	44.621	52.781
Target Compounds							
3)	L1 AR-1016-1	6.545	5.491	332.9E6	199.4E6	367.513	359.611
4)	L1 AR-1016-2	6.568	5.511	569.3E6	310.7E6	386.623	380.671
5)	L1 AR-1016-3	6.636	5.704	343.8E6	162.3E6	383.762	381.811
6)	L1 AR-1016-4	6.744	5.754	289.3E6	178.4E6	398.439	518.588 #
7)	L1 AR-1016-5	7.056	5.984	305.7E6	190.7E6	454.926	439.283
26)	L6 AR-1254-1	7.454	6.363	519.8E6	368.8E6	327.541	253.161
27)	L6 AR-1254-2	7.682	6.521	743.0E6	398.1E6	297.494	307.345
28)	L6 AR-1254-3	8.066	6.943	652.3E6	745.6E6	253.066	356.500m#
29)	L6 AR-1254-4	8.362	7.181	484.9E6	305.8E6	242.792	220.642
30)	L6 AR-1254-5	8.789	7.610	1706.9E6	1355.7E6	859.250m	698.423
31)	L7 AR-1260-1	8.231	7.078	978.7E6	699.4E6	867.485	814.513
32)	L7 AR-1260-2	8.494	7.276	1304.1E6	719.1E6	930.402	586.970 #
33)	L7 AR-1260-3	8.860	7.430	770.0E6	875.8E6	693.445	861.883
34)	L7 AR-1260-4	9.104	7.913	918.4E6	560.8E6	882.157	641.574 #
35)	L7 AR-1260-5	9.450	8.159	1854.3E6	1736.8E6	808.939	746.374

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ061721\
Data File : PQ053916.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Jun 2021 12:15
Operator : DD\AJ
Sample : M2679-06MS
Misc :
ALS Vial : 11 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 18 00:12:03 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ061621CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jun 17 06:07:32 2021
Response via : Initial Calibration
Integrator: ChemStation

Instrument :
ECD_Q
ClientSampleId :
C0CN8MS

Manual Integrations
APPROVED

Ankita
6/21/2021 10:33:22 AM

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

