

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ061722\
Data File : PQ058166.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Jun 2022 17:00
Operator : YP\AJ
Sample : N3358-18
Misc :
ALS Vial : 29 Sample Multiplier: 1


Instrument :
ECD_Q
ClientSampleId :
EW4T1

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 06/20/2022
Supervised By :Ankita Jodhani 06/20/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 25 05:31:01 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ060822CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jun 09 05:05:02 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(1) Tetrachloro-m-xylene (SA)

4.684min 4.233 ng/ml

response 62731569

(1) Tetrachloro-m-xylene #2 (SA)

4.045min 77.066 ng/ml

response 1187391155

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ061722\
Data File : PQ058166.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Jun 2022 17:00
Operator : YP\AJ
Sample : N3358-18
Misc :
ALS Vial : 29 Sample Multiplier: 1


Instrument :
ECD_Q
ClientSampleId :
EW4T1

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 06/20/2022
Supervised By :Ankita Jodhani 06/20/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 17 22:44:37 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ060822CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jun 09 05:05:02 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(1) Tetrachloro-m-xylene (SA)

4.684min 13.667 ng/ml m

response 202542553

(1) Tetrachloro-m-xylene #2 (SA)

4.045min 77.066 ng/ml

response 1187391155