

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ061819\
 Data File : PQ040952.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Jun 2019 14:59
 Operator : SM\AJ
 Sample : AR1242ICC800
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR1242401

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 19 05:41:56 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ061819CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jun 19 05:40:47 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	5.738	4.793	115.1E6	83609828	37.475	37.828
2)	SA Decachlor...	12.166	10.570	674.3E6	248.8E6	77.602	74.101
Target Compounds							
16)	L4 AR-1242-1	7.182	6.248	90578394	68989277	766.322	725.705
17)	L4 AR-1242-2	7.207	6.271	140.7E6	102.3E6	748.522	761.029
18)	L4 AR-1242-3	7.278	6.485	86838575	56497289	767.141	768.031
19)	L4 AR-1242-4	7.391	6.588	71081088	56630969	773.364	765.548
20)	L4 AR-1242-5	8.214	7.200	98622409	82765595	774.895	724.825

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ061819\
Data File : PQ040952.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Jun 2019 14:59
Operator : SM\AJ
Sample : AR1242ICC800
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1242401

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 19 05:41:56 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ061819CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jun 19 05:40:47 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

