

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ061923\
 Data File : PQ061505.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 Jun 2023 20:05
 Operator : YP\AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 144 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 19 22:20:24 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ060923.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jun 09 18:19:31 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.406	2.761	182.3E6	118.5E6	40.550	44.369
2)	SA Decachlor...	8.584	7.533	138.5E6	152.7E6	41.128	45.489
Target Compounds							
26)	L6 AR-1254-1	5.327	4.519	69576241	73889019	407.352	449.321
27)	L6 AR-1254-2	5.545	4.666	108.3E6	67491557	405.068	448.729
28)	L6 AR-1254-3	5.898	5.046	113.6E6	105.5E6	406.593	453.398
29)	L6 AR-1254-4	6.183	5.273	86471054	58972042	422.460	410.467
30)	L6 AR-1254-5	6.595	5.679	98031484	104.5E6	415.593	458.175



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ061923\
Data File : PQ061505.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 Jun 2023 20:05
Operator : YP\AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 144 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 19 22:20:24 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ060923.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jun 09 18:19:31 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.406 min
Delta R.T.: 0.000 min
Response: 182273018
Conc: 40.55 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 2.761 min
Delta R.T.: 0.000 min
Response: 118480189
Conc: 44.37 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.584 min
Delta R.T.: -0.002 min
Response: 138545491
Conc: 41.13 ng/ml



#2 Decachlorobiphenyl


R.T.: 7.533 min
Delta R.T.: -0.001 min
Response: 152685948
Conc: 45.49 ng/ml



#26 AR-1254-1


R.T.: 5.327 min
Delta R.T.: 0.000 min
Response: 69576241
Conc: 407.35 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 4.519 min
Delta R.T.: 0.000 min
Response: 73889019
Conc: 449.32 ng/ml




#27 AR-1254-2

R.T.: 5.545 min
Delta R.T.: 0.000 min
Response: 108308931
Conc: 405.07 ng/ml



#27 AR-1254-2


R.T.: 4.666 min
Delta R.T.: 0.000 min
Response: 67491557
Conc: 448.73 ng/ml



#28 AR-1254-3


R.T.: 5.898 min
Delta R.T.: -0.001 min
Response: 113619652
Conc: 406.59 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 5.046 min
Delta R.T.: -0.002 min
Response: 105505065
Conc: 453.40 ng/ml




#29 AR-1254-4

R.T.: 6.183 min
Delta R.T.: 0.000 min
Response: 86471054
Conc: 422.46 ng/ml



#29 AR-1254-4


R.T.: 5.273 min
Delta R.T.: 0.000 min
Response: 58972042
Conc: 410.47 ng/ml



#30 AR-1254-5

R.T.: 6.595 min
Delta R.T.: 0.000 min
Response: 98031484
Conc: 415.59 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 5.679 min
Delta R.T.: -0.002 min
Response: 104514303
Conc: 458.17 ng/ml