

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ062024\
 Data File : PQ067085.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Jun 2024 16:22
 Operator : YP\AJ
 Sample : AR1248ICC750
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR1248ICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 20 17:43:09 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ062024.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jun 20 17:37:39 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.452	2.825	616.7E6	453.7E6	75.423	75.536
2) SA Decachlor...	8.714	7.609	377.0E6	421.5E6	74.965	75.380
Target Compounds						
21) L5 AR-1248-1	4.563	3.833	122.2E6	93410716	751.806	746.694
22) L5 AR-1248-2	4.831	4.059	165.2E6	138.4E6	749.605	750.638
23) L5 AR-1248-3	5.024	4.096	198.7E6	133.9E6	752.777	750.507
24) L5 AR-1248-4	5.423	4.255	221.2E6	164.8E6	753.081	750.035
25) L5 AR-1248-5	5.458	4.632	221.5E6	159.2E6	753.455	749.054



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ062024\
Data File : PQ067085.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Jun 2024 16:22
Operator : YP\AJ
Sample : AR1248ICC750
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1248ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 20 17:43:09 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ062024.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jun 20 17:37:39 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.452 min
Delta R.T.: 0.000 min
Response: 616697844
Conc: 75.42 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1248ICC750




#1 Tetrachloro-m-xylene

R.T.: 2.825 min
Delta R.T.: 0.000 min
Response: 453664846
Conc: 75.54 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.714 min
Delta R.T.: 0.000 min
Response: 377042708
Conc: 74.97 ng/ml



#2 Decachlorobiphenyl


R.T.: 7.609 min
Delta R.T.: 0.000 min
Response: 421547895
Conc: 75.38 ng/ml



#21 AR-1248-1


R.T.: 4.563 min
Delta R.T.: 0.000 min
Response: 122206569
Conc: 751.81 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1248ICC750




#21 AR-1248-1

R.T.: 3.833 min
Delta R.T.: 0.000 min
Response: 93410716
Conc: 746.69 ng/ml




#22 AR-1248-2

R.T.: 4.831 min
Delta R.T.: 0.000 min
Response: 165209369
Conc: 749.60 ng/ml



#22 AR-1248-2


R.T.: 4.059 min
Delta R.T.: 0.000 min
Response: 138437816
Conc: 750.64 ng/ml



#23 AR-1248-3


R.T.: 5.024 min
Delta R.T.: 0.000 min
Response: 198709342
Conc: 752.78 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1248ICC750




#23 AR-1248-3

R.T.: 4.096 min
Delta R.T.: 0.000 min
Response: 133856303
Conc: 750.51 ng/ml




#24 AR-1248-4

R.T.: 5.423 min
Delta R.T.: 0.000 min
Response: 221223452
Conc: 753.08 ng/ml



#24 AR-1248-4


R.T.: 4.255 min
Delta R.T.: 0.000 min
Response: 164794585
Conc: 750.04 ng/ml



#25 AR-1248-5

R.T.: 5.458 min
Delta R.T.: 0.000 min
Response: 221488092
Conc: 753.45 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1248ICC750



#25 AR-1248-5

R.T.: 4.632 min
Delta R.T.: 0.000 min
Response: 159152245
Conc: 749.05 ng/ml