

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ062124\
 Data File : PQ067147.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Jun 2024 17:58
 Operator : YP\AJ
 Sample : AR1254ICV500
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 ICVPQ062124

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 21 18:12:14 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ062124.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jun 21 17:20:22 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.452	2.824	334.3E6	321.3E6	53.967	53.152
2) SA Decachlor...	8.710	7.608	213.4E6	305.6E6	56.629	50.561
Target Compounds						
26) L6 AR-1254-1	5.394	4.592	124.1E6	176.3E6	513.698	500.085
27) L6 AR-1254-2	5.614	4.740	191.0E6	155.1E6	510.633	497.431
28) L6 AR-1254-3	5.972	5.122	199.1E6	243.6E6	507.768	497.349
29) L6 AR-1254-4	6.261	5.349	139.9E6	147.8E6	510.858	498.015
30) L6 AR-1254-5	6.681	5.756	159.3E6	223.6E6	509.927	505.680



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ062124\
Data File : PQ067147.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Jun 2024 17:58
Operator : YP\AJ
Sample : AR1254ICV500
Misc :
ALS Vial : 34 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
ICVPQ062124

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 21 18:12:14 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ062124.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jun 21 17:20:22 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.452 min
Delta R.T.: 0.000 min
Response: 334335560
Conc: 53.97 ng/ml

Instrument :
ECD_Q
ClientSampleId :
ICVPQ062124




#1 Tetrachloro-m-xylene

R.T.: 2.824 min
Delta R.T.: 0.000 min
Response: 321338386
Conc: 53.15 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.710 min
Delta R.T.: -0.007 min
Response: 213390705
Conc: 56.63 ng/ml



#2 Decachlorobiphenyl


R.T.: 7.608 min
Delta R.T.: 0.000 min
Response: 305622387
Conc: 50.56 ng/ml



#26 AR-1254-1


R.T.: 5.394 min
Delta R.T.: 0.000 min
Response: 124107743
Conc: 513.70 ng/ml

Instrument :
ECD_Q
ClientSampleId :
ICVPQ062124




#26 AR-1254-1

R.T.: 4.592 min
Delta R.T.: 0.000 min
Response: 176268252
Conc: 500.09 ng/ml




#27 AR-1254-2

R.T.: 5.614 min
Delta R.T.: 0.000 min
Response: 190950169
Conc: 510.63 ng/ml



#27 AR-1254-2


R.T.: 4.740 min
Delta R.T.: 0.000 min
Response: 155058997
Conc: 497.43 ng/ml



#28 AR-1254-3


R.T.: 5.972 min
Delta R.T.: 0.000 min
Response: 199122158
Conc: 507.77 ng/ml

Instrument :
ECD_Q
ClientSampleId :
ICVPQ062124




#28 AR-1254-3

R.T.: 5.122 min
Delta R.T.: 0.000 min
Response: 243604638
Conc: 497.35 ng/ml




#29 AR-1254-4

R.T.: 6.261 min
Delta R.T.: 0.000 min
Response: 139902441
Conc: 510.86 ng/ml



#29 AR-1254-4


R.T.: 5.349 min
Delta R.T.: 0.000 min
Response: 147761286
Conc: 498.01 ng/ml



#30 AR-1254-5

R.T.: 6.681 min
Delta R.T.: 0.000 min
Response: 159311115
Conc: 509.93 ng/ml

Instrument :
ECD_Q
ClientSampleId :
ICVPQ062124



#30 AR-1254-5

R.T.: 5.756 min
Delta R.T.: 0.000 min
Response: 223638240
Conc: 505.68 ng/ml