

Method Path : \\TERASTORAGE\Terastorage\pestpcbsrv\HPCHEM1\ECD_Q\Method\
 Method File : PQ062218CLP.M
 Title : GC EXTRACTABLES
 Last Update : Fri Jun 22 04:10:40 2018
 Response Via : Initial Calibration

Calibration Files

400 =PQ029941.D 1600 =PQ029943.D 800 =PQ029942.D
 200 =PQ029940.D 100 =PQ029939.D

Compound			400	1600	800	200	100	Avg	%RSD	
1)	SA	Tetrachloro-m-xylene	1.052	0.791	0.906	1.120	1.171	1.008	E8	15.57
2)	SA	Decachlorobiphenyl	6.203	5.195	5.916	6.430	6.635	6.076	E7	9.22
3)	L1	AR-1016-1	1.611	1.127	1.348	1.706	1.805	1.519	E6	18.28
4)	L1	AR-1016-2	2.345	1.677	1.987	2.477	2.638	2.225	E6	17.48
5)	L1	AR-1016-3	2.825	2.083	2.470	3.044	3.061	2.697	E6	15.48
6)	L1	AR-1016-4	2.415	1.817	2.140	2.592	2.607	2.314	E6	14.50
7)	L1	AR-1016-5	1.882	1.419	1.677	2.016	2.049	1.809	E6	14.52
8)	L2	AR-1221-1						0.000		-1.00
9)	L2	AR-1221-2						0.000		-1.00
10)	L2	AR-1221-3						0.000		-1.00
11)	L3	AR-1232-1						0.000		-1.00
12)	L3	AR-1232-2						0.000		-1.00
13)	L3	AR-1232-3						0.000		-1.00
14)	L3	AR-1232-4						0.000		-1.00
15)	L3	AR-1232-5						0.000		-1.00
16)	L4	AR-1242-1						0.000		-1.00
17)	L4	AR-1242-2						0.000		-1.00
18)	L4	AR-1242-3						0.000		-1.00
19)	L4	AR-1242-4						0.000		-1.00
20)	L4	AR-1242-5						0.000		-1.00
21)	L5	AR-1248-1						0.000		-1.00
22)	L5	AR-1248-2						0.000		-1.00
23)	L5	AR-1248-3						0.000		-1.00
24)	L5	AR-1248-4						0.000		-1.00
25)	L5	AR-1248-5						0.000		-1.00
26)	L6	AR-1254-1						0.000		-1.00
27)	L6	AR-1254-2						0.000		-1.00
28)	L6	AR-1254-3						0.000		-1.00
29)	L6	AR-1254-4						0.000		-1.00
30)	L6	AR-1254-5						0.000		-1.00
31)	L7	AR-1260-1	3.739	2.944	3.481	4.049	4.127	3.668	E6	13.08
32)	L7	AR-1260-2	4.366	3.454	4.072	4.694	4.748	4.267	E6	12.42
33)	L7	AR-1260-3	3.127	2.493	2.938	3.375	3.444	3.075	E6	12.46
34)	L7	AR-1260-4	6.912	5.744	6.582	7.186	7.319	6.749	E6	9.31
35)	L7	AR-1260-5	4.291	3.495	4.107	4.500	4.520	4.182	E6	10.04
36)	L8	AR-1262-1						0.000		-1.00
37)	L8	AR-1262-2						0.000		-1.00
38)	L8	AR-1262-3						0.000		-1.00
39)	L8	AR-1262-4						0.000		-1.00
40)	L8	AR-1262-5						0.000		-1.00
41)	L9	AR-1268-1						0.000		-1.00
42)	L9	AR-1268-2						0.000		-1.00
43)	L9	AR-1268-3						0.000		-1.00
44)	L9	AR-1268-4						0.000		-1.00
45)	L9	AR-1268-5						0.000		-1.00

Signal #2 Calibration Files

400 =PQ029941.D 1600 =PQ029943.D 800 =PQ029942.D
 200 =PQ029940.D 100 =PQ029939.D

Compound			400	1600	800	200	100	Avg	%RSD	
1)	SA	Tetrachloro-m-xylene	1.463	1.185	1.294	1.500	1.583	1.405	E7	11.52
2)	SA	Decachlorobiphenyl	1.533	1.263	1.448	1.600	1.677	1.504	E7	10.58
3)	L1	AR-1016-1	2.460	1.822	2.073	2.658	2.854	2.373	E5	17.79
4)	L1	AR-1016-2	3.523	2.645	3.048	3.694	3.868	3.356	E5	14.93
5)	L1	AR-1016-3	4.637	3.492	4.024	4.902	5.181	4.447	E5	15.38

Method Path : \\TERASTORAGE\Terastorage\pestpcbsrv\HPCHEM1\ECD_Q\Method\
 Method File : PQ062218CLP.M
 Title : GC EXTRACTABLES
 Last Update : Fri Jun 22 04:10:40 2018
 Response Via : Initial Calibration

Calibration Files

400 =PQ029941.D 1600 =PQ029943.D 800 =PQ029942.D
 200 =PQ029940.D 100 =PQ029939.D

Compound			400	1600	800	200	100	Avg	%RSD	
6)	L1	AR-1016-4	3.743	2.863	3.295	3.924	4.150	3.595	E5	14.34
7)	L1	AR-1016-5	2.826	2.129	2.448	3.018	3.240	2.732	E5	16.30
8)	L2	AR-1221-1						0.000		-1.00
9)	L2	AR-1221-2						0.000		-1.00
10)	L2	AR-1221-3						0.000		-1.00
11)	L3	AR-1232-1						0.000		-1.00
12)	L3	AR-1232-2						0.000		-1.00
13)	L3	AR-1232-3						0.000		-1.00
14)	L3	AR-1232-4						0.000		-1.00
15)	L3	AR-1232-5						0.000		-1.00
16)	L4	AR-1242-1						0.000		-1.00
17)	L4	AR-1242-2						0.000		-1.00
18)	L4	AR-1242-3						0.000		-1.00
19)	L4	AR-1242-4						0.000		-1.00
20)	L4	AR-1242-5						0.000		-1.00
21)	L5	AR-1248-1						0.000		-1.00
22)	L5	AR-1248-2						0.000		-1.00
23)	L5	AR-1248-3						0.000		-1.00
24)	L5	AR-1248-4						0.000		-1.00
25)	L5	AR-1248-5						0.000		-1.00
26)	L6	AR-1254-1						0.000		-1.00
27)	L6	AR-1254-2						0.000		-1.00
28)	L6	AR-1254-3						0.000		-1.00
29)	L6	AR-1254-4						0.000		-1.00
30)	L6	AR-1254-5						0.000		-1.00
31)	L7	AR-1260-1	7.566	5.857	6.834	8.083	8.701	7.408	E5	14.92
32)	L7	AR-1260-2	0.954	0.742	0.868	1.022	1.078	0.933	E6	14.18
33)	L7	AR-1260-3	8.709	6.912	8.037	9.170	9.501	8.466	E5	12.14
34)	L7	AR-1260-4	1.860	1.511	1.747	1.937	1.982	1.808	E6	10.41
35)	L7	AR-1260-5	1.310	1.076	1.240	1.367	1.401	1.279	E6	10.10
36)	L8	AR-1262-1						0.000		-1.00
37)	L8	AR-1262-2						0.000		-1.00
38)	L8	AR-1262-3						0.000		-1.00
39)	L8	AR-1262-4						0.000		-1.00
40)	L8	AR-1262-5						0.000		-1.00
41)	L9	AR-1268-1						0.000		-1.00
42)	L9	AR-1268-2						0.000		-1.00
43)	L9	AR-1268-3						0.000		-1.00
44)	L9	AR-1268-4						0.000		-1.00
45)	L9	AR-1268-5						0.000		-1.00

(#) = Out of Range