

Method Path : \\TERASTORAGE\Terastorage\pestpcbsrv\HPCHEM1\ECD_Q\Method\
 Method File : PQ062218CLP.M
 Title : GC EXTRACTABLES
 Last Update : Fri Jun 22 05:02:15 2018
 Response Via : Initial Calibration

Calibration Files

400 =PQ029941.D 1600 =PQ029943.D 800 =PQ029942.D
 200 =PQ029940.D 100 =PQ029939.D

Compound			400	1600	800	200	100	Avg	%RSD	
1)	SA	Tetrachloro-m-xylene	1.052	0.791	0.906	1.120	1.171	1.008	E8	15.57
2)	SA	Decachlorobiphenyl	6.203	5.195	5.916	6.430	6.635	6.076	E7	9.22
3)	L1	AR-1016-1	1.611	1.127	1.348	1.706	1.805	1.519	E6	18.28
4)	L1	AR-1016-2	2.345	1.677	1.987	2.477	2.638	2.225	E6	17.48
5)	L1	AR-1016-3	2.825	2.083	2.470	3.044	3.061	2.697	E6	15.48
6)	L1	AR-1016-4	2.415	1.817	2.140	2.592	2.607	2.314	E6	14.50
7)	L1	AR-1016-5	1.882	1.419	1.677	2.016	2.049	1.809	E6	14.52
8)	L2	AR-1221-1					1.174	1.174	E6	0.00
9)	L2	AR-1221-2					8.061	8.061	E5	0.00
10)	L2	AR-1221-3					2.792	2.792	E6	0.00
11)	L3	AR-1232-1					2.101	2.101	E6	0.00
12)	L3	AR-1232-2					1.089	1.089	E6	0.00
13)	L3	AR-1232-3					1.150	1.150	E6	0.00
14)	L3	AR-1232-4					9.978	9.978	E5	0.00
15)	L3	AR-1232-5					1.004	1.004	E6	0.00
16)	L4	AR-1242-1					2.348	2.348	E6	0.00
17)	L4	AR-1242-2					4.931	4.931	E6	0.00
18)	L4	AR-1242-3					2.365	2.365	E6	0.00
19)	L4	AR-1242-4					1.787	1.787	E6	0.00
20)	L4	AR-1242-5					2.512	2.512	E6	0.00
21)	L5	AR-1248-1	3.418	2.964	3.230	3.410	3.748	3.354	E6	8.56
22)	L5	AR-1248-2	3.046	2.653	2.893	3.025	3.277	2.979	E6	7.68
23)	L5	AR-1248-3	3.877	3.362	3.692	3.907	4.207	3.809	E6	8.16
24)	L5	AR-1248-4	3.710	3.256	3.523	3.671	4.006	3.633	E6	7.54
25)	L5	AR-1248-5	3.794	3.286	3.575	3.745	4.110	3.702	E6	8.16
26)	L6	AR-1254-1	3.257	2.873	3.174	3.748	3.978	3.406	E6	13.16
27)	L6	AR-1254-2	4.955	4.422	4.812	5.613	5.769	5.114	E6	11.03
28)	L6	AR-1254-3	5.279	4.819	5.188	5.850	6.064	5.440	E6	9.34
29)	L6	AR-1254-4	3.878	3.550	3.796	4.317	4.486	4.005	E6	9.64
30)	L6	AR-1254-5	2.933	2.711	2.897	3.221	3.260	3.004	E6	7.71
31)	L7	AR-1260-1	3.739	2.944	3.481	4.049	4.127	3.668	E6	13.08
32)	L7	AR-1260-2	4.366	3.454	4.072	4.694	4.748	4.267	E6	12.42
33)	L7	AR-1260-3	3.127	2.493	2.938	3.375	3.444	3.075	E6	12.46
34)	L7	AR-1260-4	6.912	5.744	6.582	7.186	7.319	6.749	E6	9.31
35)	L7	AR-1260-5	4.291	3.495	4.107	4.500	4.520	4.182	E6	10.04
36)	L8	AR-1262-1					2.661	2.661	E6	0.00
37)	L8	AR-1262-2					3.610	3.610	E6	0.00
38)	L8	AR-1262-3					5.377	5.377	E6	0.00
39)	L8	AR-1262-4					2.242	2.242	E6	0.00
40)	L8	AR-1262-5					4.629	4.629	E6	0.00
41)	L9	AR-1268-1					8.088	8.088	E6	0.00
42)	L9	AR-1268-2					7.063	7.063	E6	0.00
43)	L9	AR-1268-3					1.663	1.663	E6	0.00
44)	L9	AR-1268-4					3.042	3.042	E6	0.00
45)	L9	AR-1268-5					2.240	2.240	E7	0.00

Signal #2 Calibration Files

400 =PQ029941.D 1600 =PQ029943.D 800 =PQ029942.D
 200 =PQ029940.D 100 =PQ029939.D

Compound		400	1600	800	200	100	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	1.463	1.185	1.294	1.500	1.583	1.405	E7	11.52
2) SA	Decachlorobiphenyl	1.533	1.263	1.448	1.600	1.677	1.504	E7	10.58
3) L1	AR-1016-1	2.460	1.822	2.073	2.658	2.854	2.373	E5	17.79
4) L1	AR-1016-2	3.523	2.645	3.048	3.694	3.868	3.356	E5	14.93
5) L1	AR-1016-3	4.637	3.492	4.024	4.902	5.181	4.447	E5	15.38

Method Path : \\TERASTORAGE\Terastorage\pestpcbsrv\HPCHEM1\ECD_Q\Method\
 Method File : PQ062218CLP.M
 Title : GC EXTRACTABLES
 Last Update : Fri Jun 22 05:02:15 2018
 Response Via : Initial Calibration

Calibration Files

400 =PQ029941.D 1600 =PQ029943.D 800 =PQ029942.D
 200 =PQ029940.D 100 =PQ029939.D

	Compound	400	1600	800	200	100	Avg		%RSD
6)	L1 AR-1016-4	3.743	2.863	3.295	3.924	4.150	3.595	E5	14.34
7)	L1 AR-1016-5	2.826	2.129	2.448	3.018	3.240	2.732	E5	16.30
8)	L2 AR-1221-1					1.891	1.891	E5	0.00
9)	L2 AR-1221-2					1.314	1.314	E5	0.00
10)	L2 AR-1221-3					4.482	4.482	E5	0.00
11)	L3 AR-1232-1					3.295	3.295	E5	0.00
12)	L3 AR-1232-2					1.730	1.730	E5	0.00
13)	L3 AR-1232-3					1.601	1.601	E5	0.00
14)	L3 AR-1232-4					1.992	1.992	E5	0.00
15)	L3 AR-1232-5					1.716	1.716	E5	0.00
16)	L4 AR-1242-1					3.573	3.573	E5	0.00
17)	L4 AR-1242-2					7.444	7.444	E5	0.00
18)	L4 AR-1242-3					3.680	3.680	E5	0.00
19)	L4 AR-1242-4					2.607	2.607	E5	0.00
20)	L4 AR-1242-5					2.945	2.945	E5	0.00
21)	L5 AR-1248-1	5.757	4.988	5.347	5.819	6.603	5.703	E5	10.61
22)	L5 AR-1248-2	3.461	3.060	3.241	3.552	4.026	3.468	E5	10.56
23)	L5 AR-1248-3	5.994	5.288	5.613	6.155	6.995	6.009	E5	10.76
24)	L5 AR-1248-4	2.404	2.198	2.300	2.588	2.635	2.425	E5	7.66
25)	L5 AR-1248-5	1.373	1.278	1.339	1.499	1.437	1.385	E5	6.20
26)	L6 AR-1254-1	8.199	7.522	8.078	9.451	9.421	8.534	E5	10.10
27)	L6 AR-1254-2	6.958	6.400	6.874	8.118	7.918	7.254	E5	10.11
28)	L6 AR-1254-3	1.184	1.103	1.173	1.329	1.310	1.220	E6	7.89
29)	L6 AR-1254-4	8.052	7.438	7.877	9.045	8.999	8.282	E5	8.59
30)	L6 AR-1254-5	5.561	5.259	5.524	6.119	5.989	5.690	E5	6.23
31)	L7 AR-1260-1	7.566	5.857	6.834	8.083	8.701	7.408	E5	14.92
32)	L7 AR-1260-2	0.954	0.742	0.868	1.022	1.078	0.933	E6	14.18
33)	L7 AR-1260-3	8.709	6.912	8.037	9.170	9.501	8.466	E5	12.14
34)	L7 AR-1260-4	1.860	1.511	1.747	1.937	1.982	1.808	E6	10.41
35)	L7 AR-1260-5	1.310	1.076	1.240	1.367	1.401	1.279	E6	10.10
36)	L8 AR-1262-1					4.956	4.956	E5	0.00
37)	L8 AR-1262-2					1.326	1.326	E6	0.00
38)	L8 AR-1262-3					9.874	9.874	E5	0.00
39)	L8 AR-1262-4					8.844	8.844	E5	0.00
40)	L8 AR-1262-5					3.402	3.402	E5	0.00
41)	L9 AR-1268-1					2.560	2.560	E6	0.00
42)	L9 AR-1268-2					2.184	2.184	E6	0.00
43)	L9 AR-1268-3					5.580	5.580	E5	0.00
44)	L9 AR-1268-4					9.196	9.196	E5	0.00
45)	L9 AR-1268-5					6.866	6.866	E6	0.00

(#) = Out of Range