

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ062222\
 Data File : PQ058297.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Jun 2022 22:34
 Operator : YP\AJ
 Sample : AR1262ICC100
 Misc :
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR12621016

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 06/23/2022
 Supervised By :Ankita Jodhani 06/23/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 23 10:05:37 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ062222CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jun 23 09:42:21 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.685	4.050	77009994	52916675	5.283	4.927
2) SA Decachlor...	10.604	9.192	153.4E6	84915616	10.657	10.219
Target Compounds						
36) L8 AR-1262-1	8.093	7.167	71696484	31482549	105.094	97.793
37) L8 AR-1262-2	8.674	7.704	147.4E6	113.5E6	102.458	96.661
38) L8 AR-1262-3	8.999	7.991	74984587	43931633	102.139m	99.738
39) L8 AR-1262-4	9.099	8.055	61206013	82104491	104.687m	98.197
40) L8 AR-1262-5	9.806	8.581	64803485	36858551	105.039	99.120

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ062222\
Data File : PQ058297.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Jun 2022 22:34
Operator : YP\AJ
Sample : AR1262ICC100
Misc :
ALS Vial : 33 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR12621016

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/23/2022
Supervised By :Ankita Jodhani 06/23/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 23 10:05:37 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ062222CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jun 23 09:42:21 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

