

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ062419\
 Data File : PQ041070.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Jun 2019 17:50
 Operator : SM\AJ
 Sample : AR1221ICC400
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 25 05:33:45 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ062419CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jun 25 05:26:29 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml



System Monitoring Compounds							
1)	SA Tetrachlo...	5.719	4.774	142.5E6	106.4E6	20.133	20.000
2)	SA Decachlor...	12.145	10.552	440.5E6	174.9E6	40.000	40.000
Target Compounds							
8)	L2 AR-1221-1	5.997	5.084	28935824	24089100	400.000	400.000
9)	L2 AR-1221-2	6.107	5.202	21645764	17378624	400.000	400.000
10)	L2 AR-1221-3	6.203	5.306	72818591	59039647	400.000	400.000


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ062419\
Data File : PQ041070.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Jun 2019 17:50
Operator : SM\AJ
Sample : AR1221ICC400
Misc :
ALS Vial : 10 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 25 05:33:45 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ062419CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jun 25 05:26:29 2019
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 5.719 min
Delta R.T.: 0.000 min
Response: 142548357
Conc: 20.13 ng/ml




#1 Tetrachloro-m-xylene

R.T.: 4.774 min
Delta R.T.: 0.000 min
Response: 106401316
Conc: 20.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 12.145 min
Delta R.T.: 0.000 min
Response: 440467970
Conc: 40.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.552 min
Delta R.T.: 0.000 min
Response: 174905917
Conc: 40.00 ng/ml




#8 AR-1221-1

R.T.: 5.997 min
Delta R.T.: 0.000 min
Response: 28935824
Conc: 400.00 ng/ml




#8 AR-1221-1

R.T.: 5.084 min
Delta R.T.: 0.000 min
Response: 24089100
Conc: 400.00 ng/ml




#9 AR-1221-2

R.T.: 6.107 min
Delta R.T.: 0.000 min
Response: 21645764
Conc: 400.00 ng/ml




#9 AR-1221-2

R.T.: 5.202 min
Delta R.T.: 0.000 min
Response: 17378624
Conc: 400.00 ng/ml



#10 AR-1221-3

R.T.: 6.203 min
Delta R.T.: 0.000 min
Response: 72818591
Conc: 400.00 ng/ml



#10 AR-1221-3

R.T.: 5.306 min
Delta R.T.: 0.000 min
Response: 59039647
Conc: 400.00 ng/ml