

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ062525\
 Data File : PQ070415.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Jun 2025 20:03
 Operator : YP\AJ
 Sample : Q2126-07
 Misc : AR1242 LOD 40 PPB
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 LOD-MDL-WATER-01-QT2-2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 25 07:32:42 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ062425.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jun 25 02:08:47 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.403	2.782	106.2E6	119.8E6	17.349	17.744
2)	SA Decachlor...	8.500	7.542	80505349	140.8E6	19.557	19.422
Target Compounds							
3)	L1 AR-1016-1	4.484	3.784	5770191	6295373	30.742	30.313
4)	L1 AR-1016-2	4.503	3.799	8862530	9644978	28.619	28.515
5)	L1 AR-1016-3	4.560	3.959	5770131	5196606	28.679	27.693
6)	L1 AR-1016-4	4.650	4.007	4415340	4773080	27.895	27.208
7)	L1 AR-1016-5	4.935	4.203	4294646	5453462	26.407	27.019
8)	L2 AR-1221-1	3.578	2.982	2978273	705659	34.737	6.902 #
9)	L2 AR-1221-2	3.676	3.052	2307558	2813546	35.703	35.668
10)	L2 AR-1221-3	3.743	3.126	3576113	3851906	18.438	16.994
11)	L3 AR-1232-1	3.743	3.126	3576113	3851906	25.272	23.108
12)	L3 AR-1232-2	4.230	3.799	4796584	9644978	59.192	60.446
13)	L3 AR-1232-3	4.503	3.959	8862530	5196606	62.108	59.125
14)	L3 AR-1232-4	4.650	4.045	4415340	5379051	61.407	69.724
15)	L3 AR-1232-5	4.788	4.203	4185926	5453462	70.041	65.122
16)	L4 AR-1242-1	4.484	3.784	5770191	6295373	37.285	35.884
17)	L4 AR-1242-2	4.503	3.799	8862530	9644978	33.738	34.637
18)	L4 AR-1242-3	4.560	3.959	5770131	5196606	33.964	33.065
19)	L4 AR-1242-4	4.650	4.045	4415340	5379051	32.717	35.322
20)	L4 AR-1242-5	5.357	4.540	4944558	6816011	36.765	37.354
21)	L5 AR-1248-1	4.484	3.784	5770191	6295373	46.591	44.788
22)	L5 AR-1248-2	4.747	4.007	3561334	4773080	19.304	20.446
23)	L5 AR-1248-3	4.935	4.045	4294646	5379051	19.448	23.894
24)	L5 AR-1248-4	5.322	4.203	4765513	5453462	21.474	19.999
25)	L5 AR-1248-5	5.357	4.578	4944558	6537544	21.395	26.105
26)	L6 AR-1254-1	5.295	4.540	3394233	6816011	13.721	16.824
27)	L6 AR-1254-2	5.506	4.687	4065549	2365345	10.824	6.550 #
28)	L6 AR-1254-3	5.858	5.071	2072666	6065382	5.338	10.725 #
29)	L6 AR-1254-4	6.159	5.293	3555980	1872727	14.285	6.140 #
30)	L6 AR-1254-5	6.535	5.698	2207107	540945	7.430	1.051 #
31)	L7 AR-1260-1	6.018	5.200	397568	1757410	1.376	4.601 #
32)	L7 AR-1260-2	6.280	5.378	177624	690474	0.557	1.576 #
33)	L7 AR-1260-3	6.620	5.545	92392	24170839	0.356	56.022 #
34)	L7 AR-1260-4	6.865	5.980	82092	226122	0.301	0.616 #
35)	L7 AR-1260-5	7.156	6.229	135064	604667	0.271	0.727 #
36)	L8 AR-1262-1	6.865	5.737	82092	522749	0.233	0.955 #
37)	L8 AR-1262-2	7.156	5.997	135064	273865	0.238	0.539 #
38)	L8 AR-1262-3	7.457f	0.000	541663	0	1.390	N.D. #
39)	L8 AR-1262-4	7.496	6.619f	146536	59223	0.458	0.085 #
40)	L8 AR-1262-5	8.031f	7.072	53966	227302	0.288	0.717 #
41)	L9 AR-1268-1	7.457f	0.000	541663	0	0.752	N.D. #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ062525\
 Data File : PQ070415.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Jun 2025 20:03
 Operator : YP\AJ
 Sample : Q2126-07
 Misc : AR1242 LOD 40 PPB
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 LOD-MDL-WATER-01-QT2-2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 25 07:32:42 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ062425.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jun 25 02:08:47 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	7.496	6.619f	146536	59223	0.221	0.055 #
43) L9	AR-1268-3	7.683	0.000	150771	0	0.266	N.D. #
44) L9	AR-1268-4	8.031f	7.072	53966	227302	0.262	0.604 #
45) L9	AR-1268-5	8.279	7.324	106530	369734	0.067	0.141 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ062525\
Data File : PQ070415.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Jun 2025 20:03
Operator : YP\AJ
Sample : Q2126-07
Misc : AR1242 LOD 40 PPB
ALS Vial : 16 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT2-2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 25 07:32:42 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ062425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jun 25 02:08:47 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.403 min
Delta R.T.: 0.000 min
Response: 106208501
Conc: 17.35 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT2-2025




#1 Tetrachloro-m-xylene

R.T.: 2.782 min
Delta R.T.: 0.000 min
Response: 119829557
Conc: 17.74 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.500 min
Delta R.T.: -0.002 min
Response: 80505349
Conc: 19.56 ng/ml



#2 Decachlorobiphenyl


R.T.: 7.542 min
Delta R.T.: 0.000 min
Response: 140798247
Conc: 19.42 ng/ml



#3 AR-1016-1


R.T.: 4.484 min
Delta R.T.: -0.001 min
Response: 5770191
Conc: 30.74 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT2-2025




#3 AR-1016-1

R.T.: 3.784 min
Delta R.T.: 0.000 min
Response: 6295373
Conc: 30.31 ng/ml




#4 AR-1016-2

R.T.: 4.503 min
Delta R.T.: 0.000 min
Response: 8862530
Conc: 28.62 ng/ml



#4 AR-1016-2


R.T.: 3.799 min
Delta R.T.: 0.000 min
Response: 9644978
Conc: 28.52 ng/ml



#5 AR-1016-3


R.T.: 4.560 min
Delta R.T.: -0.001 min
Response: 5770131
Conc: 28.68 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT2-2025




#5 AR-1016-3

R.T.: 3.959 min
Delta R.T.: 0.000 min
Response: 5196606
Conc: 27.69 ng/ml




#6 AR-1016-4

R.T.: 4.650 min
Delta R.T.: 0.000 min
Response: 4415340
Conc: 27.89 ng/ml



#6 AR-1016-4


R.T.: 4.007 min
Delta R.T.: 0.000 min
Response: 4773080
Conc: 27.21 ng/ml



#7 AR-1016-5


R.T.: 4.935 min
Delta R.T.: 0.000 min
Response: 4294646
Conc: 26.41 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT2-2025




#7 AR-1016-5

R.T.: 4.203 min
Delta R.T.: 0.000 min
Response: 5453462
Conc: 27.02 ng/ml




#8 AR-1221-1

R.T.: 3.578 min
Delta R.T.: -0.019 min
Response: 2978273
Conc: 34.74 ng/ml



#8 AR-1221-1


R.T.: 2.982 min
Delta R.T.: 0.000 min
Response: 705659
Conc: 6.90 ng/ml



#9 AR-1221-2


R.T.: 3.676 min
Delta R.T.: 0.000 min
Response: 2307558
Conc: 35.70 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT2-2025




#9 AR-1221-2

R.T.: 3.052 min
Delta R.T.: -0.006 min
Response: 2813546
Conc: 35.67 ng/ml




#10 AR-1221-3

R.T.: 3.743 min
Delta R.T.: 0.000 min
Response: 3576113
Conc: 18.44 ng/ml



#10 AR-1221-3


R.T.: 3.126 min
Delta R.T.: -0.001 min
Response: 3851906
Conc: 16.99 ng/ml



#11 AR-1232-1


R.T.: 3.743 min
Delta R.T.: 0.000 min
Response: 3576113
Conc: 25.27 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT2-2025




#11 AR-1232-1

R.T.: 3.126 min
Delta R.T.: 0.000 min
Response: 3851906
Conc: 23.11 ng/ml




#12 AR-1232-2

R.T.: 4.230 min
Delta R.T.: 0.000 min
Response: 4796584
Conc: 59.19 ng/ml



#12 AR-1232-2


R.T.: 3.799 min
Delta R.T.: 0.000 min
Response: 9644978
Conc: 60.45 ng/ml



#13 AR-1232-3


R.T.: 4.503 min
Delta R.T.: 0.000 min
Response: 8862530
Conc: 62.11 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT2-2025




#13 AR-1232-3

R.T.: 3.959 min
Delta R.T.: 0.000 min
Response: 5196606
Conc: 59.12 ng/ml




#14 AR-1232-4

R.T.: 4.650 min
Delta R.T.: 0.000 min
Response: 4415340
Conc: 61.41 ng/ml



#14 AR-1232-4


R.T.: 4.045 min
Delta R.T.: 0.000 min
Response: 5379051
Conc: 69.72 ng/ml



#15 AR-1232-5


R.T.: 4.788 min
Delta R.T.: 0.000 min
Response: 4185926
Conc: 70.04 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT2-2025




#15 AR-1232-5

R.T.: 4.203 min
Delta R.T.: 0.000 min
Response: 5453462
Conc: 65.12 ng/ml




#16 AR-1242-1

R.T.: 4.484 min
Delta R.T.: 0.000 min
Response: 5770191
Conc: 37.28 ng/ml



#16 AR-1242-1


R.T.: 3.784 min
Delta R.T.: 0.000 min
Response: 6295373
Conc: 35.88 ng/ml



#17 AR-1242-2


R.T.: 4.503 min
Delta R.T.: 0.000 min
Response: 8862530
Conc: 33.74 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT2-2025




#17 AR-1242-2

R.T.: 3.799 min
Delta R.T.: 0.000 min
Response: 9644978
Conc: 34.64 ng/ml




#18 AR-1242-3

R.T.: 4.560 min
Delta R.T.: 0.000 min
Response: 5770131
Conc: 33.96 ng/ml



#18 AR-1242-3


R.T.: 3.959 min
Delta R.T.: 0.000 min
Response: 5196606
Conc: 33.06 ng/ml



#19 AR-1242-4


R.T.: 4.650 min
Delta R.T.: 0.000 min
Response: 4415340
Conc: 32.72 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT2-2025




#19 AR-1242-4

R.T.: 4.045 min
Delta R.T.: 0.000 min
Response: 5379051
Conc: 35.32 ng/ml




#20 AR-1242-5

R.T.: 5.357 min
Delta R.T.: 0.000 min
Response: 4944558
Conc: 36.76 ng/ml



#20 AR-1242-5


R.T.: 4.540 min
Delta R.T.: 0.000 min
Response: 6816011
Conc: 37.35 ng/ml



#21 AR-1248-1


R.T.: 4.484 min
Delta R.T.: 0.000 min
Response: 5770191
Conc: 46.59 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT2-2025




#21 AR-1248-1

R.T.: 3.784 min
Delta R.T.: 0.000 min
Response: 6295373
Conc: 44.79 ng/ml




#22 AR-1248-2

R.T.: 4.747 min
Delta R.T.: 0.000 min
Response: 3561334
Conc: 19.30 ng/ml



#22 AR-1248-2


R.T.: 4.007 min
Delta R.T.: -0.001 min
Response: 4773080
Conc: 20.45 ng/ml



#23 AR-1248-3


R.T.: 4.935 min
Delta R.T.: 0.000 min
Response: 4294646
Conc: 19.45 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT2-2025




#23 AR-1248-3

R.T.: 4.045 min
Delta R.T.: 0.000 min
Response: 5379051
Conc: 23.89 ng/ml




#24 AR-1248-4

R.T.: 5.322 min
Delta R.T.: 0.000 min
Response: 4765513
Conc: 21.47 ng/ml



#24 AR-1248-4


R.T.: 4.203 min
Delta R.T.: 0.000 min
Response: 5453462
Conc: 20.00 ng/ml



#25 AR-1248-5


R.T.: 5.357 min
Delta R.T.: 0.000 min
Response: 4944558
Conc: 21.39 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT2-2025




#25 AR-1248-5

R.T.: 4.578 min
Delta R.T.: 0.000 min
Response: 6537544
Conc: 26.10 ng/ml




#26 AR-1254-1

R.T.: 5.295 min
Delta R.T.: -0.002 min
Response: 3394233
Conc: 13.72 ng/ml



#26 AR-1254-1


R.T.: 4.540 min
Delta R.T.: 0.000 min
Response: 6816011
Conc: 16.82 ng/ml



#27 AR-1254-2


R.T.: 5.506 min
Delta R.T.: -0.005 min
Response: 4065549
Conc: 10.82 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT2-2025




#27 AR-1254-2

R.T.: 4.687 min
Delta R.T.: 0.000 min
Response: 2365345
Conc: 6.55 ng/ml




#28 AR-1254-3

R.T.: 5.858 min
Delta R.T.: 0.000 min
Response: 2072666
Conc: 5.34 ng/ml



#28 AR-1254-3


R.T.: 5.071 min
Delta R.T.: 0.003 min
Response: 6065382
Conc: 10.72 ng/ml



#29 AR-1254-4


R.T.: 6.159 min
Delta R.T.: 0.020 min
Response: 3555980
Conc: 14.29 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT2-2025




#29 AR-1254-4

R.T.: 5.293 min
Delta R.T.: -0.001 min
Response: 1872727
Conc: 6.14 ng/ml




#30 AR-1254-5

R.T.: 6.535 min
Delta R.T.: -0.013 min
Response: 2207107
Conc: 7.43 ng/ml



#30 AR-1254-5


R.T.: 5.698 min
Delta R.T.: 0.000 min
Response: 540945
Conc: 1.05 ng/ml



#31 AR-1260-1


R.T.: 6.018 min
Delta R.T.: -0.007 min
Response: 397568
Conc: 1.38 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT2-2025




#31 AR-1260-1

R.T.: 5.200 min
Delta R.T.: 0.003 min
Response: 1757410
Conc: 4.60 ng/ml




#32 AR-1260-2

R.T.: 6.280 min
Delta R.T.: 0.000 min
Response: 177624
Conc: 0.56 ng/ml



#32 AR-1260-2


R.T.: 5.378 min
Delta R.T.: -0.009 min
Response: 690474
Conc: 1.58 ng/ml



#33 AR-1260-3


R.T.: 6.620 min
Delta R.T.: -0.011 min
Response: 92392
Conc: 0.36 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT2-2025




#33 AR-1260-3

R.T.: 5.545 min
Delta R.T.: 0.018 min
Response: 24170839
Conc: 56.02 ng/ml




#34 AR-1260-4

R.T.: 6.865 min
Delta R.T.: 0.019 min
Response: 82092
Conc: 0.30 ng/ml



#34 AR-1260-4


R.T.: 5.980 min
Delta R.T.: -0.008 min
Response: 226122
Conc: 0.62 ng/ml



#35 AR-1260-5


R.T.: 7.156 min
Delta R.T.: 0.005 min
Response: 135064
Conc: 0.27 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT2-2025




#35 AR-1260-5

R.T.: 6.229 min
Delta R.T.: -0.003 min
Response: 604667
Conc: 0.73 ng/ml




#36 AR-1262-1

R.T.: 6.865 min
Delta R.T.: 0.019 min
Response: 82092
Conc: 0.23 ng/ml



#36 AR-1262-1


R.T.: 5.737 min
Delta R.T.: -0.004 min
Response: 522749
Conc: 0.96 ng/ml



#37 AR-1262-2


R.T.: 7.156 min
Delta R.T.: 0.005 min
Response: 135064
Conc: 0.24 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT2-2025




#37 AR-1262-2

R.T.: 5.997 min
Delta R.T.: 0.009 min
Response: 273865
Conc: 0.54 ng/ml




#38 AR-1262-3

R.T.: 7.457 min
Delta R.T.: 0.033 min
Response: 541663
Conc: 1.39 ng/ml



#38 AR-1262-3


R.T.: 0.000 min
Exp R.T. : 6.510 min
Response: 0
Conc: N.D.



#39 AR-1262-4


R.T.: 7.496 min
Delta R.T.: -0.002 min
Response: 146536
Conc: 0.46 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT2-2025




#39 AR-1262-4

R.T.: 6.619 min
Delta R.T.: 0.047 min
Response: 59223
Conc: 0.08 ng/ml




#40 AR-1262-5

R.T.: 8.031 min
Delta R.T.: 0.038 min
Response: 53966
Conc: 0.29 ng/ml



#40 AR-1262-5


R.T.: 7.072 min
Delta R.T.: 0.012 min
Response: 227302
Conc: 0.72 ng/ml



#41 AR-1268-1


R.T.: 7.457 min
Delta R.T.: 0.033 min
Response: 541663
Conc: 0.75 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT2-2025




#41 AR-1268-1

R.T.: 0.000 min
Exp R.T. : 6.509 min
Response: 0
Conc: N.D.




#42 AR-1268-2

R.T.: 7.496 min
Delta R.T.: -0.004 min
Response: 146536
Conc: 0.22 ng/ml



#42 AR-1268-2


R.T.: 6.619 min
Delta R.T.: 0.047 min
Response: 59223
Conc: 0.06 ng/ml



#43 AR-1268-3


R.T.: 7.683 min
Delta R.T.: 0.001 min
Response: 150771
Conc: 0.27 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT2-2025




#43 AR-1268-3

R.T.: 0.000 min
Exp R.T. : 6.773 min
Response: 0
Conc: N.D.




#44 AR-1268-4

R.T.: 8.031 min
Delta R.T.: 0.037 min
Response: 53966
Conc: 0.26 ng/ml



#44 AR-1268-4


R.T.: 7.072 min
Delta R.T.: 0.012 min
Response: 227302
Conc: 0.60 ng/ml



#45 AR-1268-5

R.T.: 8.279 min
Delta R.T.: 0.001 min
Response: 106530
Conc: 0.07 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT2-2025



#45 AR-1268-5

R.T.: 7.324 min
Delta R.T.: -0.004 min
Response: 369734
Conc: 0.14 ng/ml