

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ062525\
 Data File : PQ070463.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 Jun 2025 12:06
 Operator : YP\AJ
 Sample : Q2126-09
 Misc : AR1268 MDL 25 PPB
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 MDL-WATER-03-QT2-2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 26 04:57:32 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ062425.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jun 26 02:49:53 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.403	2.781	110.7E6	123.0E6	18.090	18.214
2)	SA Decachlor...	8.503	7.542	92241295	158.7E6	22.408	21.889
Target Compounds							
3)	L1 AR-1016-1	4.478	3.785	233369	259988	1.243	1.252
4)	L1 AR-1016-2	4.492	3.798	356995	271734	1.153	0.803 #
5)	L1 AR-1016-3	4.562	3.966	211106	254240	1.049	1.355 #
6)	L1 AR-1016-4	4.646	4.006	328123	364965	2.073	2.080
7)	L1 AR-1016-5	4.938	4.200	121277	168977	0.746	0.837
8)	L2 AR-1221-1	3.617	2.992	327251	179929	3.817	1.760 #
9)	L2 AR-1221-2	3.676	3.050	2368314	2347326	36.643	29.758
10)	L2 AR-1221-3	3.751	3.115	243216	235331	1.254	1.038
11)	L3 AR-1232-1	3.751	3.115	243216	235331	1.719	1.412
12)	L3 AR-1232-2	4.229	3.798	168981	271734	2.085	1.703
13)	L3 AR-1232-3	4.492	3.966	356995	254240	2.502	2.893
14)	L3 AR-1232-4	4.646	4.040	328123	378508	4.563	4.906
15)	L3 AR-1232-5	4.792	4.200	129285	168977	2.163	2.018
16)	L4 AR-1242-1	4.478	3.785	233369	259988	1.508	1.482
17)	L4 AR-1242-2	4.492	3.798	356995	271734	1.359	0.976 #
18)	L4 AR-1242-3	4.562	3.966	211106	254240	1.243	1.618 #
19)	L4 AR-1242-4	4.646	4.040	328123	378508	2.431	2.486
20)	L4 AR-1242-5	5.354	4.545	629378	396973	4.680	2.176 #
21)	L5 AR-1248-1	4.478	3.785	233369	259988	1.884	1.850
22)	L5 AR-1248-2	4.744	4.006	105858	364965	0.574	1.563 #
23)	L5 AR-1248-3	4.938	4.040	121277	378508	0.549	1.681 #
24)	L5 AR-1248-4	5.326	4.200	2080755	168977	9.376	0.620 #
25)	L5 AR-1248-5	5.354	4.584	629378	481175	2.723	1.921 #
26)	L6 AR-1254-1	5.293	4.545	1327874	396973	5.368	0.980 #
27)	L6 AR-1254-2	5.506	4.692	1172104	308897	3.120	0.855 #
28)	L6 AR-1254-3	5.854	5.077	272044	3754638	0.701	6.639 #
29)	L6 AR-1254-4	6.161	5.297	2465802	123241	9.906	0.404 #
30)	L6 AR-1254-5	6.571	5.739f	262601	5556051	0.884	10.796 #
31)	L7 AR-1260-1	6.027	5.174	123556	162147	0.428	0.424
32)	L7 AR-1260-2	6.313f	5.385	177604	27533	0.557	0.063 #
33)	L7 AR-1260-3	6.632	5.545	2682563	1329092	10.330	3.081 #
34)	L7 AR-1260-4	6.840	5.986	70099	6871131	0.257	18.724 #
35)	L7 AR-1260-5	7.150	6.231	1566305	2712991	3.144	3.264
36)	L8 AR-1262-1	6.840	5.739	70099	5556051	0.199	10.154 #
37)	L8 AR-1262-2	7.150	5.986	1566305	6871131	2.765	13.511 #
38)	L8 AR-1262-3	7.423	6.507	13584863	22717277	34.873	57.863 #
39)	L8 AR-1262-4	7.500	6.570	12456498	20375800	38.930	29.086 #
40)	L8 AR-1262-5	7.993	7.058	4047110	7020867	21.603	22.140
41)	L9 AR-1268-1	7.423	6.507	13584863	22717277	18.851	19.602

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ062525\
 Data File : PQ070463.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 Jun 2025 12:06
 Operator : YP\AJ
 Sample : Q2126-09
 Misc : AR1268 MDL 25 PPB
 ALS Vial : 29 Sample Multiplier: 1


Instrument :
 ECD_Q
 ClientSampleId :
 MDL-WATER-03-QT2-2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 26 04:57:32 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ062425.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jun 26 02:49:53 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	7.500	6.570	12456498	20375800	18.753	19.003
43) L9	AR-1268-3	7.682	6.771	10715404	19584314	18.894	20.516
44) L9	AR-1268-4	7.993	7.058	4047110	7020867	19.618	18.666
45) L9	AR-1268-5	8.277	7.325	31073958	52926935	19.409	20.138


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.



#1 Tetrachloro-m-xylene


R.T.: 3.403 min
Delta R.T.: 0.000 min
Response: 110741221
Conc: 18.09 ng/ml

Instrument :
ECD_Q
ClientSampleId :
MDL-WATER-03-QT2-2025




#1 Tetrachloro-m-xylene

R.T.: 2.781 min
Delta R.T.: 0.000 min
Response: 122999551
Conc: 18.21 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.503 min
Delta R.T.: 0.000 min
Response: 92241295
Conc: 22.41 ng/ml



#2 Decachlorobiphenyl


R.T.: 7.542 min
Delta R.T.: -0.001 min
Response: 158681806
Conc: 21.89 ng/ml



#3 AR-1016-1


R.T.: 4.478 min
Delta R.T.: -0.006 min
Response: 233369
Conc: 1.24 ng/ml

Instrument :
ECD_Q
ClientSampleId :
MDL-WATER-03-QT2-2025




#3 AR-1016-1

R.T.: 3.785 min
Delta R.T.: 0.001 min
Response: 259988
Conc: 1.25 ng/ml




#4 AR-1016-2

R.T.: 4.492 min
Delta R.T.: -0.011 min
Response: 356995
Conc: 1.15 ng/ml



#4 AR-1016-2


R.T.: 3.798 min
Delta R.T.: -0.002 min
Response: 271734
Conc: 0.80 ng/ml



#5 AR-1016-3


R.T.: 4.562 min
Delta R.T.: 0.002 min
Response: 211106
Conc: 1.05 ng/ml

Instrument :
ECD_Q
ClientSampleId :
MDL-WATER-03-QT2-2025




#5 AR-1016-3

R.T.: 3.966 min
Delta R.T.: 0.006 min
Response: 254240
Conc: 1.35 ng/ml




#6 AR-1016-4

R.T.: 4.646 min
Delta R.T.: -0.004 min
Response: 328123
Conc: 2.07 ng/ml



#6 AR-1016-4


R.T.: 4.006 min
Delta R.T.: -0.001 min
Response: 364965
Conc: 2.08 ng/ml



#7 AR-1016-5


R.T.: 4.938 min
Delta R.T.: 0.003 min
Response: 121277
Conc: 0.75 ng/ml

Instrument :
ECD_Q
ClientSampleId :
MDL-WATER-03-QT2-2025




#7 AR-1016-5

R.T.: 4.200 min
Delta R.T.: -0.003 min
Response: 168977
Conc: 0.84 ng/ml




#8 AR-1221-1

R.T.: 3.617 min
Delta R.T.: 0.019 min
Response: 327251
Conc: 3.82 ng/ml



#8 AR-1221-1


R.T.: 2.992 min
Delta R.T.: 0.010 min
Response: 179929
Conc: 1.76 ng/ml



#9 AR-1221-2


R.T.: 3.676 min
Delta R.T.: 0.000 min
Response: 2368314
Conc: 36.64 ng/ml

Instrument :
ECD_Q
ClientSampleId :
MDL-WATER-03-QT2-2025




#9 AR-1221-2

R.T.: 3.050 min
Delta R.T.: -0.009 min
Response: 2347326
Conc: 29.76 ng/ml




#10 AR-1221-3

R.T.: 3.751 min
Delta R.T.: 0.009 min
Response: 243216
Conc: 1.25 ng/ml



#10 AR-1221-3


R.T.: 3.115 min
Delta R.T.: -0.012 min
Response: 235331
Conc: 1.04 ng/ml



#11 AR-1232-1


R.T.: 3.751 min
Delta R.T.: 0.008 min
Response: 243216
Conc: 1.72 ng/ml

Instrument :
ECD_Q
ClientSampleId :
MDL-WATER-03-QT2-2025




#11 AR-1232-1

R.T.: 3.115 min
Delta R.T.: -0.012 min
Response: 235331
Conc: 1.41 ng/ml




#12 AR-1232-2

R.T.: 4.229 min
Delta R.T.: -0.001 min
Response: 168981
Conc: 2.09 ng/ml



#12 AR-1232-2


R.T.: 3.798 min
Delta R.T.: -0.002 min
Response: 271734
Conc: 1.70 ng/ml



#13 AR-1232-3


R.T.: 4.492 min
Delta R.T.: -0.012 min
Response: 356995
Conc: 2.50 ng/ml

Instrument :
ECD_Q
ClientSampleId :
MDL-WATER-03-QT2-2025




#13 AR-1232-3

R.T.: 3.966 min
Delta R.T.: 0.007 min
Response: 254240
Conc: 2.89 ng/ml




#14 AR-1232-4

R.T.: 4.646 min
Delta R.T.: -0.005 min
Response: 328123
Conc: 4.56 ng/ml



#14 AR-1232-4


R.T.: 4.040 min
Delta R.T.: -0.005 min
Response: 378508
Conc: 4.91 ng/ml



#15 AR-1232-5


R.T.: 4.792 min
Delta R.T.: 0.003 min
Response: 129285
Conc: 2.16 ng/ml

Instrument :
ECD_Q
ClientSampleId :
MDL-WATER-03-QT2-2025




#15 AR-1232-5

R.T.: 4.200 min
Delta R.T.: -0.003 min
Response: 168977
Conc: 2.02 ng/ml




#16 AR-1242-1

R.T.: 4.478 min
Delta R.T.: -0.006 min
Response: 233369
Conc: 1.51 ng/ml



#16 AR-1242-1


R.T.: 3.785 min
Delta R.T.: 0.001 min
Response: 259988
Conc: 1.48 ng/ml



#17 AR-1242-2


R.T.: 4.492 min
Delta R.T.: -0.011 min
Response: 356995
Conc: 1.36 ng/ml

Instrument :
ECD_Q
ClientSampleId :
MDL-WATER-03-QT2-2025




#17 AR-1242-2

R.T.: 3.798 min
Delta R.T.: -0.002 min
Response: 271734
Conc: 0.98 ng/ml




#18 AR-1242-3

R.T.: 4.562 min
Delta R.T.: 0.002 min
Response: 211106
Conc: 1.24 ng/ml



#18 AR-1242-3


R.T.: 3.966 min
Delta R.T.: 0.007 min
Response: 254240
Conc: 1.62 ng/ml



#19 AR-1242-4


R.T.: 4.646 min
Delta R.T.: -0.005 min
Response: 328123
Conc: 2.43 ng/ml

Instrument :
ECD_Q
ClientSampleId :
MDL-WATER-03-QT2-2025




#19 AR-1242-4

R.T.: 4.040 min
Delta R.T.: -0.005 min
Response: 378508
Conc: 2.49 ng/ml




#20 AR-1242-5

R.T.: 5.354 min
Delta R.T.: -0.003 min
Response: 629378
Conc: 4.68 ng/ml



#20 AR-1242-5


R.T.: 4.545 min
Delta R.T.: 0.005 min
Response: 396973
Conc: 2.18 ng/ml



#21 AR-1248-1


R.T.: 4.478 min
Delta R.T.: -0.006 min
Response: 233369
Conc: 1.88 ng/ml

Instrument :
ECD_Q
ClientSampleId :
MDL-WATER-03-QT2-2025




#21 AR-1248-1

R.T.: 3.785 min
Delta R.T.: 0.000 min
Response: 259988
Conc: 1.85 ng/ml




#22 AR-1248-2

R.T.: 4.744 min
Delta R.T.: -0.004 min
Response: 105858
Conc: 0.57 ng/ml



#22 AR-1248-2


R.T.: 4.006 min
Delta R.T.: -0.002 min
Response: 364965
Conc: 1.56 ng/ml



#23 AR-1248-3


R.T.: 4.938 min
Delta R.T.: 0.003 min
Response: 121277
Conc: 0.55 ng/ml

Instrument :
ECD_Q
ClientSampleId :
MDL-WATER-03-QT2-2025




#23 AR-1248-3

R.T.: 4.040 min
Delta R.T.: -0.005 min
Response: 378508
Conc: 1.68 ng/ml




#24 AR-1248-4

R.T.: 5.326 min
Delta R.T.: 0.005 min
Response: 2080755
Conc: 9.38 ng/ml



#24 AR-1248-4


R.T.: 4.200 min
Delta R.T.: -0.004 min
Response: 168977
Conc: 0.62 ng/ml



#25 AR-1248-5


R.T.: 5.354 min
Delta R.T.: -0.003 min
Response: 629378
Conc: 2.72 ng/ml

Instrument :
ECD_Q
ClientSampleId :
MDL-WATER-03-QT2-2025




#25 AR-1248-5

R.T.: 4.584 min
Delta R.T.: 0.006 min
Response: 481175
Conc: 1.92 ng/ml




#26 AR-1254-1

R.T.: 5.293 min
Delta R.T.: -0.003 min
Response: 1327874
Conc: 5.37 ng/ml



#26 AR-1254-1


R.T.: 4.545 min
Delta R.T.: 0.005 min
Response: 396973
Conc: 0.98 ng/ml



#27 AR-1254-2


R.T.: 5.506 min
Delta R.T.: -0.005 min
Response: 1172104
Conc: 3.12 ng/ml

Instrument :
ECD_Q
ClientSampleId :
MDL-WATER-03-QT2-2025




#27 AR-1254-2

R.T.: 4.692 min
Delta R.T.: 0.006 min
Response: 308897
Conc: 0.86 ng/ml




#28 AR-1254-3

R.T.: 5.854 min
Delta R.T.: -0.004 min
Response: 272044
Conc: 0.70 ng/ml



#28 AR-1254-3


R.T.: 5.077 min
Delta R.T.: 0.009 min
Response: 3754638
Conc: 6.64 ng/ml



#29 AR-1254-4


R.T.: 6.161 min
Delta R.T.: 0.021 min
Response: 2465802
Conc: 9.91 ng/ml

Instrument :
ECD_Q
ClientSampleId :
MDL-WATER-03-QT2-2025




#29 AR-1254-4

R.T.: 5.297 min
Delta R.T.: 0.003 min
Response: 123241
Conc: 0.40 ng/ml




#30 AR-1254-5

R.T.: 6.571 min
Delta R.T.: 0.023 min
Response: 262601
Conc: 0.88 ng/ml



#30 AR-1254-5


R.T.: 5.739 min
Delta R.T.: 0.041 min
Response: 5556051
Conc: 10.80 ng/ml



#31 AR-1260-1


R.T.: 6.027 min
Delta R.T.: 0.003 min
Response: 123556
Conc: 0.43 ng/ml

Instrument :
ECD_Q
ClientSampleId :
MDL-WATER-03-QT2-2025




#31 AR-1260-1

R.T.: 5.174 min
Delta R.T.: -0.022 min
Response: 162147
Conc: 0.42 ng/ml




#32 AR-1260-2

R.T.: 6.313 min
Delta R.T.: 0.032 min
Response: 177604
Conc: 0.56 ng/ml



#32 AR-1260-2


R.T.: 5.385 min
Delta R.T.: 0.000 min
Response: 27533
Conc: 0.06 ng/ml



#33 AR-1260-3


R.T.: 6.632 min
Delta R.T.: 0.001 min
Response: 2682563
Conc: 10.33 ng/ml

Instrument :
ECD_Q
ClientSampleId :
MDL-WATER-03-QT2-2025




#33 AR-1260-3

R.T.: 5.545 min
Delta R.T.: 0.017 min
Response: 1329092
Conc: 3.08 ng/ml




#34 AR-1260-4

R.T.: 6.840 min
Delta R.T.: -0.005 min
Response: 70099
Conc: 0.26 ng/ml



#34 AR-1260-4


R.T.: 5.986 min
Delta R.T.: -0.001 min
Response: 6871131
Conc: 18.72 ng/ml



#35 AR-1260-5


R.T.: 7.150 min
Delta R.T.: 0.000 min
Response: 1566305
Conc: 3.14 ng/ml

Instrument :
ECD_Q
ClientSampleId :
MDL-WATER-03-QT2-2025




#35 AR-1260-5

R.T.: 6.231 min
Delta R.T.: 0.000 min
Response: 2712991
Conc: 3.26 ng/ml




#36 AR-1262-1

R.T.: 6.840 min
Delta R.T.: -0.005 min
Response: 70099
Conc: 0.20 ng/ml



#36 AR-1262-1


R.T.: 5.739 min
Delta R.T.: -0.003 min
Response: 5556051
Conc: 10.15 ng/ml



#37 AR-1262-2


R.T.: 7.150 min
Delta R.T.: -0.001 min
Response: 1566305
Conc: 2.77 ng/ml

Instrument :
ECD_Q
ClientSampleId :
MDL-WATER-03-QT2-2025




#37 AR-1262-2

R.T.: 5.986 min
Delta R.T.: -0.002 min
Response: 6871131
Conc: 13.51 ng/ml




#38 AR-1262-3

R.T.: 7.423 min
Delta R.T.: -0.001 min
Response: 13584863
Conc: 34.87 ng/ml



#38 AR-1262-3


R.T.: 6.507 min
Delta R.T.: -0.003 min
Response: 22717277
Conc: 57.86 ng/ml



#39 AR-1262-4


R.T.: 7.500 min
Delta R.T.: 0.002 min
Response: 12456498
Conc: 38.93 ng/ml

Instrument :
ECD_Q
ClientSampleId :
MDL-WATER-03-QT2-2025




#39 AR-1262-4

R.T.: 6.570 min
Delta R.T.: -0.001 min
Response: 20375800
Conc: 29.09 ng/ml




#40 AR-1262-5

R.T.: 7.993 min
Delta R.T.: 0.000 min
Response: 4047110
Conc: 21.60 ng/ml



#40 AR-1262-5


R.T.: 7.058 min
Delta R.T.: -0.002 min
Response: 7020867
Conc: 22.14 ng/ml



#41 AR-1268-1


R.T.: 7.423 min
Delta R.T.: 0.000 min
Response: 13584863
Conc: 18.85 ng/ml

Instrument :
ECD_Q
ClientSampleId :
MDL-WATER-03-QT2-2025




#41 AR-1268-1

R.T.: 6.507 min
Delta R.T.: -0.002 min
Response: 22717277
Conc: 19.60 ng/ml




#42 AR-1268-2

R.T.: 7.500 min
Delta R.T.: 0.000 min
Response: 12456498
Conc: 18.75 ng/ml



#42 AR-1268-2


R.T.: 6.570 min
Delta R.T.: -0.001 min
Response: 20375800
Conc: 19.00 ng/ml



#43 AR-1268-3


R.T.: 7.682 min
Delta R.T.: 0.000 min
Response: 10715404
Conc: 18.89 ng/ml

Instrument :
ECD_Q
ClientSampleId :
MDL-WATER-03-QT2-2025




#43 AR-1268-3

R.T.: 6.771 min
Delta R.T.: -0.002 min
Response: 19584314
Conc: 20.52 ng/ml




#44 AR-1268-4

R.T.: 7.993 min
Delta R.T.: 0.000 min
Response: 4047110
Conc: 19.62 ng/ml



#44 AR-1268-4


R.T.: 7.058 min
Delta R.T.: -0.002 min
Response: 7020867
Conc: 18.67 ng/ml



#45 AR-1268-5

R.T.: 8.277 min
Delta R.T.: 0.000 min
Response: 31073958
Conc: 19.41 ng/ml

Instrument :
ECD_Q
ClientSampleId :
MDL-WATER-03-QT2-2025



#45 AR-1268-5

R.T.: 7.325 min
Delta R.T.: -0.003 min
Response: 52926935
Conc: 20.14 ng/ml