

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ062625\
 Data File : PQ070483.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 Jun 2025 18:55
 Operator : YP\AJ
 Sample : Q2126-08
 Misc : AR1232 LOQ 50 PPB
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 LOQ-WATER-02-QT2-2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 26 05:30:30 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ062425.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jun 26 02:49:53 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.404	2.782	99422851	112.9E6	16.241	16.724
2) SA Decachlor...	8.501	7.542	81558823	139.4E6	19.813	19.227
Target Compounds						
11) L3 AR-1232-1	3.743	3.126	5876454	6192664	41.529	37.151
12) L3 AR-1232-2	4.229	3.798	3434563	6555389	42.384	41.083
13) L3 AR-1232-3	4.503	3.959	5748776	3663070	40.287	41.677
14) L3 AR-1232-4	4.650	4.044	2589911	3132736	36.019	40.607
15) L3 AR-1232-5	4.789	4.203	1694590	3063332	28.355	36.580 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ062625\
Data File : PQ070483.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 Jun 2025 18:55
Operator : YP\AJ
Sample : Q2126-08
Misc : AR1232 LOQ 50 PPB
ALS Vial : 16 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
LOQ-WATER-02-QT2-2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 26 05:30:30 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ062425.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jun 26 02:49:53 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

