

Method Path : \\TERASTORAGE\Terastorage\pestpcbsrv\HPCHEM1\ECD_Q\Method\
 Method File : PQ062918.M
 Title : GC EXTRACTABLES
 Last Update : Fri Jun 29 01:53:51 2018
 Response Via : Initial Calibration

Calibration Files

500 =PQ030204.D 1000 =PQ030202.D 750 =PQ030203.D
 250 =PQ030205.D 50 =PQ030206.D

Compound		500	1000	750	250	50	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	4.068	3.925	4.008	4.351	4.636	4.197	E7	6.98
2) SA	Decachlorobiphenyl	3.532	3.317	3.376	3.590	3.851	3.533	E7	5.93
3) L1	AR-1016-1	6.858	6.370	6.732	7.524	7.442	6.985	E5	7.00
4) L1	AR-1016-2	0.935	0.878	0.929	1.015	0.947	0.941	E6	5.23
5) L1	AR-1016-3	1.297	1.194	1.271	1.389	1.419	1.314	E6	6.92
6) L1	AR-1016-4	1.093	1.013	1.076	1.158	1.174	1.103	E6	5.91
7) L1	AR-1016-5	1.126	1.048	1.103	1.196	1.026	1.100	E6	6.10
8) L2	AR-1221-1	5.613					5.613	E5	0.00
9) L2	AR-1221-2	3.734					3.734	E5	0.00
10) L2	AR-1221-3	1.119					1.119	E6	0.00
11) L3	AR-1232-1	7.365					7.365	E5	0.00
12) L3	AR-1232-2	3.767					3.767	E5	0.00
13) L3	AR-1232-3	5.715					5.715	E5	0.00
14) L3	AR-1232-4	4.542					4.542	E5	0.00
15) L3	AR-1232-5	3.867					3.867	E5	0.00
16) L4	AR-1242-1	7.628					7.628	E5	0.00
17) L4	AR-1242-2	1.159					1.159	E6	0.00
18) L4	AR-1242-3	1.035					1.035	E6	0.00
19) L4	AR-1242-4	8.536					8.536	E5	0.00
20) L4	AR-1242-5	9.307					9.307	E5	0.00
21) L5	AR-1248-1	1.507					1.507	E6	0.00
22) L5	AR-1248-2	1.260					1.260	E6	0.00
23) L5	AR-1248-3	2.026					2.026	E6	0.00
24) L5	AR-1248-4	1.614					1.614	E6	0.00
25) L5	AR-1248-5	1.688					1.688	E6	0.00
26) L6	AR-1254-1	2.182					2.182	E6	0.00
27) L6	AR-1254-2	2.588					2.588	E6	0.00
28) L6	AR-1254-3	2.809					2.809	E6	0.00
29) L6	AR-1254-4	1.991					1.991	E6	0.00
30) L6	AR-1254-5	2.479					2.479	E6	0.00
31) L7	AR-1260-1	1.998	1.899	1.993	1.990	2.149	2.006	E6	4.50
32) L7	AR-1260-2	2.331	2.227	2.272	2.425	2.612	2.373	E6	6.43
33) L7	AR-1260-3	2.857	2.722	2.770	2.918	2.833	2.820	E6	2.70
34) L7	AR-1260-4	1.984	1.907	1.936	2.037	2.142	2.001	E6	4.63
35) L7	AR-1260-5	3.993	3.862	3.948	4.011	4.110	3.985	E6	2.28
36) L8	AR-1262-1	1.426					1.426	E6	0.00
37) L8	AR-1262-2	1.835					1.835	E6	0.00
38) L8	AR-1262-3	2.297					2.297	E6	0.00
39) L8	AR-1262-4	2.204					2.204	E6	0.00
40) L8	AR-1262-5	5.002					5.002	E6	0.00
41) L9	AR-1268-1	4.236					4.236	E6	0.00
42) L9	AR-1268-2	3.865					3.865	E6	0.00
43) L9	AR-1268-3	3.255					3.255	E6	0.00
44) L9	AR-1268-4	1.546					1.546	E6	0.00
45) L9	AR-1268-5	1.050					1.050	E7	0.00

Signal #2 Calibration Files

500 =PQ030204.D 1000 =PQ030202.D 750 =PQ030203.D
 250 =PQ030205.D 50 =PQ030206.D

Compound		500	1000	750	250	50	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	1.405	1.382	1.419	1.375	1.301	1.376	E7	3.30
2) SA	Decachlorobiphenyl	2.051	1.978	2.016	2.084	2.079	2.042	E7	2.20
3) L1	AR-1016-1	3.735	3.645	3.775	3.836	3.935	3.785	E5	2.87
4) L1	AR-1016-2	3.913	3.755	3.889	3.901	3.874	3.866	E5	1.66
5) L1	AR-1016-3	3.038	2.881	2.988	3.049	3.267	3.045	E5	4.64

Method Path : \\TERASTORAGE\Terastorage\pestpcbsrv\HPCHEM1\ECD_Q\Method\
 Method File : PQ062918.M
 Title : GC EXTRACTABLES
 Last Update : Fri Jun 29 01:53:51 2018
 Response Via : Initial Calibration

Calibration Files

500 =PQ030204.D 1000 =PQ030202.D 750 =PQ030203.D
 250 =PQ030205.D 50 =PQ030206.D

Compound			500	1000	750	250	50	Avg		%RSD

6)	L1	AR-1016-4	4.013	3.805	3.965	4.096	4.549	4.086	E5	6.85
7)	L1	AR-1016-5	4.499	4.116	4.333	4.353	3.499	4.160	E5	9.48
8)	L2	AR-1221-1	1.681					1.681	E5	0.00
9)	L2	AR-1221-2	1.223					1.223	E5	0.00
10)	L2	AR-1221-3	3.868					3.868	E5	0.00
11)	L3	AR-1232-1	2.443					2.443	E5	0.00
12)	L3	AR-1232-2	2.153					2.153	E5	0.00
13)	L3	AR-1232-3	1.405					1.405	E5	0.00
14)	L3	AR-1232-4	1.244					1.244	E5	0.00
15)	L3	AR-1232-5	1.924					1.924	E5	0.00
16)	L4	AR-1242-1	2.856					2.856	E5	0.00
17)	L4	AR-1242-2	6.066					6.066	E5	0.00
18)	L4	AR-1242-3	2.989					2.989	E5	0.00
19)	L4	AR-1242-4	3.103					3.103	E5	0.00
20)	L4	AR-1242-5	3.238					3.238	E5	0.00
21)	L5	AR-1248-1	4.019					4.019	E5	0.00
22)	L5	AR-1248-2	4.436					4.436	E5	0.00
23)	L5	AR-1248-3	4.968					4.968	E5	0.00
24)	L5	AR-1248-4	8.361					8.361	E5	0.00
25)	L5	AR-1248-5	8.361					8.361	E5	0.00
26)	L6	AR-1254-1	8.948					8.948	E5	0.00
27)	L6	AR-1254-2	7.503					7.503	E5	0.00
28)	L6	AR-1254-3	1.271					1.271	E6	0.00
29)	L6	AR-1254-4	8.768					8.768	E5	0.00
30)	L6	AR-1254-5	5.291					5.291	E5	0.00
31)	L7	AR-1260-1	0.994	0.949	0.981	1.002	1.048	0.995	E6	3.61
32)	L7	AR-1260-2	1.158	1.116	1.155	1.163	1.144	1.147	E6	1.65
33)	L7	AR-1260-3	9.475	9.110	9.412	9.413	9.214	9.325	E5	1.67
34)	L7	AR-1260-4	2.484	2.405	2.470	2.472	2.323	2.431	E6	2.79
35)	L7	AR-1260-5	1.772	1.718	1.768	1.771	1.728	1.751	E6	1.49
36)	L8	AR-1262-1	7.644					7.644	E5	0.00
37)	L8	AR-1262-2	1.013					1.013	E6	0.00
38)	L8	AR-1262-3	1.259					1.259	E6	0.00
39)	L8	AR-1262-4	1.080					1.080	E6	0.00
40)	L8	AR-1262-5	2.422					2.422	E6	0.00
41)	L9	AR-1268-1	2.389					2.389	E6	0.00
42)	L9	AR-1268-2	2.414					2.414	E6	0.00
43)	L9	AR-1268-3	1.956					1.956	E6	0.00
44)	L9	AR-1268-4	8.382					8.382	E5	0.00
45)	L9	AR-1268-5	6.356					6.356	E6	0.00

(#) = Out of Range