

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_Q\Data\PQ062920\  
 Data File : PQ048527.D  
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch  
 Acq On : 29 Jun 2020 10:42  
 Operator : DD\AJ  
 Sample : I.BLK  
 Misc :  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 ECD\_Q  
 ClientSampleId :  
 I.BLK

Integration File signal 1: autoint1.e  
 Integration File signal 2: autoint2.e  
 Quant Time: Jun 29 13:32:38 2020  
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_Q\Method\PQ061720.M  
 Quant Title : GC EXTRACTABLES  
 QLast Update : Wed Jun 17 15:37:20 2020  
 Response via : Initial Calibration  
 Integrator: ChemStation

Volume Inj. : 2 µl  
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2  
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

| Compound                    | RT#1   | RT#2  | Resp#1  | Resp#2   | ng/ml  | ng/ml  |
|-----------------------------|--------|-------|---------|----------|--------|--------|
| -----                       |        |       |         |          |        |        |
| System Monitoring Compounds |        |       |         |          |        |        |
| 1) SA Tetrachlo...          | 5.290  | 4.286 | 110.9E6 | 56859723 | 15.044 | 16.105 |
| 2) SA Decachlor...          | 11.562 | 9.659 | 144.6E6 | 85305730 | 14.943 | 17.727 |

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_Q\Data\PQ062920\  
Data File : PQ048527.D  
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch  
Acq On : 29 Jun 2020 10:42  
Operator : DD\AJ  
Sample : I.BLK  
Misc :  
ALS Vial : 2 Sample Multiplier: 1

Instrument :  
ECD\_Q  
ClientSampleId :  
I.BLK

Integration File signal 1: autoint1.e  
Integration File signal 2: autoint2.e  
Quant Time: Jun 29 13:32:38 2020  
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_Q\Method\PQ061720.M  
Quant Title : GC EXTRACTABLES  
QLast Update : Wed Jun 17 15:37:20 2020  
Response via : Initial Calibration  
Integrator: ChemStation

Volume Inj. : 2 µl  
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2  
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

