

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ063022\
 Data File : PQ058492.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 Jun 2022 17:00
 Operator : YP\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 01 00:02:25 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ062922.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jun 29 16:14:42 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.688	4.043	565.5E6	493.4E6	44.176	45.528
2) SA Decachlor...	10.608	9.179	583.4E6	379.8E6	48.484	50.649
Target Compounds						
3) L1 AR-1016-1	5.873	5.146	141.0E6	159.9E6	437.269	457.817
4) L1 AR-1016-2	5.896	5.166	246.5E6	228.8E6	437.847	456.055
5) L1 AR-1016-3	5.958	5.347	158.8E6	115.3E6	438.054	457.096
6) L1 AR-1016-4	6.059	5.385	134.7E6	91393513	442.233	458.066
7) L1 AR-1016-5	6.353	5.603	109.7E6	117.2E6	452.220	450.945
31) L7 AR-1260-1	7.479	6.642	180.6E6	220.2E6	448.366	468.016
32) L7 AR-1260-2	7.736	6.826	216.4E6	265.0E6	447.946	470.170
33) L7 AR-1260-3	8.024	6.985	289.4E6	247.1E6	450.173	474.396
34) L7 AR-1260-4	8.335	7.458	200.5E6	185.3E6	455.378	482.168
35) L7 AR-1260-5	8.675	7.695	454.1E6	452.2E6	458.345	486.524

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ063022\
Data File : PQ058492.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Jun 2022 17:00
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 01 00:02:25 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ062922.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jun 29 16:14:42 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

