

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ070123\
 Data File : PQ061852.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Jul 2023 07:12
 Operator : YP\AJ
 Sample : AR1254ICC1000
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 01 12:30:43 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ070123.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Jul 01 12:19:21 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.404	2.759	492.1E6	256.8E6	95.277	93.880
2)	SA Decachlor...	8.583	7.528	351.3E6	322.4E6	91.903	93.308
Target Compounds							
26)	L6 AR-1254-1	5.327	4.516	174.9E6	151.4E6	925.623	915.510
27)	L6 AR-1254-2	5.544	4.662	272.9E6	137.8E6	924.431	906.562
28)	L6 AR-1254-3	5.897	5.044	291.9E6	222.3E6	930.823	933.969
29)	L6 AR-1254-4	6.182	5.270	213.6E6	120.5E6	928.012	845.620
30)	L6 AR-1254-5	6.595	5.675	246.0E6	221.8E6	935.302	937.877



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ070123\
Data File : PQ061852.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Jul 2023 07:12
Operator : YP\AJ
Sample : AR1254ICC1000
Misc :
ALS Vial : 53 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 01 12:30:43 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ070123.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Jul 01 12:19:21 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.404 min
Delta R.T.: 0.000 min
Response: 492136204
Conc: 95.28 ng/ml

Instrument :
ECD_Q
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 2.759 min
Delta R.T.: 0.000 min
Response: 256807396
Conc: 93.88 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.583 min
Delta R.T.: 0.000 min
Response: 351330623
Conc: 91.90 ng/ml



#2 Decachlorobiphenyl


R.T.: 7.528 min
Delta R.T.: 0.000 min
Response: 322443740
Conc: 93.31 ng/ml



#26 AR-1254-1


R.T.: 5.327 min
Delta R.T.: 0.000 min
Response: 174947235
Conc: 925.62 ng/ml

Instrument :
ECD_Q
ClientSampleId :




#26 AR-1254-1

R.T.: 4.516 min
Delta R.T.: 0.000 min
Response: 151435820
Conc: 915.51 ng/ml




#27 AR-1254-2

R.T.: 5.544 min
Delta R.T.: 0.000 min
Response: 272892699
Conc: 924.43 ng/ml



#27 AR-1254-2


R.T.: 4.662 min
Delta R.T.: 0.000 min
Response: 137765424
Conc: 906.56 ng/ml



#28 AR-1254-3


R.T.: 5.897 min
Delta R.T.: 0.000 min
Response: 291946447
Conc: 930.82 ng/ml

Instrument :
ECD_Q
ClientSampleId :




#28 AR-1254-3

R.T.: 5.044 min
Delta R.T.: 0.000 min
Response: 222279918
Conc: 933.97 ng/ml




#29 AR-1254-4

R.T.: 6.182 min
Delta R.T.: 0.000 min
Response: 213632998
Conc: 928.01 ng/ml



#29 AR-1254-4


R.T.: 5.270 min
Delta R.T.: 0.000 min
Response: 120542981
Conc: 845.62 ng/ml



#30 AR-1254-5

R.T.: 6.595 min
Delta R.T.: 0.001 min
Response: 246047511
Conc: 935.30 ng/ml

Instrument :
ECD_Q
ClientSampleId :



#30 AR-1254-5

R.T.: 5.675 min
Delta R.T.: 0.000 min
Response: 221751688
Conc: 937.88 ng/ml