

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ070123\
 Data File : PQ061853.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Jul 2023 07:27
 Operator : YP\AJ
 Sample : AR1254ICC750
 Misc :
 ALS Vial : 54 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 01 12:31:05 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ070123.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Jul 01 12:19:21 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.404	2.758	379.5E6	200.4E6	73.468	73.248
2)	SA Decachlor...	8.583	7.528	275.1E6	251.1E6	71.966	72.672
Target Compounds							
26)	L6 AR-1254-1	5.327	4.516	136.7E6	119.4E6	723.081	721.648
27)	L6 AR-1254-2	5.544	4.662	213.0E6	108.2E6	721.667	712.294
28)	L6 AR-1254-3	5.897	5.043	226.9E6	172.7E6	723.546	725.616
29)	L6 AR-1254-4	6.182	5.270	166.6E6	96543059	723.795	677.258
30)	L6 AR-1254-5	6.595	5.676	191.9E6	171.7E6	729.559	726.302



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ070123\
Data File : PQ061853.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Jul 2023 07:27
Operator : YP\AJ
Sample : AR1254ICC750
Misc :
ALS Vial : 54 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 01 12:31:05 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ070123.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Jul 01 12:19:21 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.404 min
Delta R.T.: 0.000 min
Response: 379483153
Conc: 73.47 ng/ml

Instrument :
ECD_Q
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 2.758 min
Delta R.T.: 0.000 min
Response: 200367854
Conc: 73.25 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.583 min
Delta R.T.: 0.000 min
Response: 275112608
Conc: 71.97 ng/ml



#2 Decachlorobiphenyl


R.T.: 7.528 min
Delta R.T.: 0.000 min
Response: 251133217
Conc: 72.67 ng/ml



#26 AR-1254-1


R.T.: 5.327 min
Delta R.T.: 0.000 min
Response: 136665864
Conc: 723.08 ng/ml

Instrument :
ECD_Q
ClientSampleId :




#26 AR-1254-1

R.T.: 4.516 min
Delta R.T.: 0.000 min
Response: 119368911
Conc: 721.65 ng/ml




#27 AR-1254-2

R.T.: 5.544 min
Delta R.T.: 0.000 min
Response: 213036695
Conc: 721.67 ng/ml



#27 AR-1254-2


R.T.: 4.662 min
Delta R.T.: 0.000 min
Response: 108243537
Conc: 712.29 ng/ml



#28 AR-1254-3


R.T.: 5.897 min
Delta R.T.: 0.000 min
Response: 226935347
Conc: 723.55 ng/ml

Instrument :
ECD_Q
ClientSampleId :




#28 AR-1254-3

R.T.: 5.043 min
Delta R.T.: 0.000 min
Response: 172692903
Conc: 725.62 ng/ml




#29 AR-1254-4

R.T.: 6.182 min
Delta R.T.: 0.000 min
Response: 166621079
Conc: 723.79 ng/ml



#29 AR-1254-4


R.T.: 5.270 min
Delta R.T.: 0.000 min
Response: 96543059
Conc: 677.26 ng/ml



#30 AR-1254-5

R.T.: 6.595 min
Delta R.T.: 0.001 min
Response: 191923354
Conc: 729.56 ng/ml

Instrument :
ECD_Q
ClientSampleId :



#30 AR-1254-5

R.T.: 5.676 min
Delta R.T.: 0.000 min
Response: 171726973
Conc: 726.30 ng/ml