

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ070618\
 Data File : PQ030400.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 Jul 2018 12:00
 Operator : UA/SM
 Sample : J3773-01
 Misc :
 ALS Vial : 142 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 EPE-119-5(30-32)

Manual Integrations
 APPROVED

mohammad
 7/6/2018 12:28:29 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 05 12:15:09 2018
 Quant Method : \\TERASTORAGE\Terastorage\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ062918.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jun 29 01:53:51 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.432	3.702	707.3E6	240.2E6	16.852	17.456m
2) SA Decachlor...	10.109	8.670	346.7E6	223.9E6	9.814	10.965

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ070618\
Data File : PQ030400.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 05 Jul 2018 12:00
Operator : UA/SM
Sample : J3773-01
Misc :
ALS Vial : 142 Sample Multiplier: 1



Instrument :
ECD_Q
ClientSampleId :
EPE-119-5(30-32)


Manual Integrations
APPROVED

mohammad
7/6/2018 12:28:29 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 05 12:15:09 2018
Quant Method : \\TERASTORAGE\Terastorage\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ062918.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jun 29 01:53:51 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.432 min
Delta R.T.: -0.005 min
Response: 707333325
Conc: 16.85 ng/ml

Instrument :
ECD_Q
ClientSampleId :
EPE-119-5(30-32)


Manual Integrations
APPROVED

mohammad
7/6/2018 12:28:29 PM




#1 Tetrachloro-m-xylene

R.T.: 3.702 min
Delta R.T.: -0.001 min
Response: 240233473
Conc: 17.46 ng/ml m



#2 Decachlorobiphenyl

R.T.: 10.109 min
Delta R.T.: -0.010 min
Response: 346749804
Conc: 9.81 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.670 min
Delta R.T.: -0.006 min
Response: 223869423
Conc: 10.97 ng/ml