

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ070618\
 Data File : PQ030407.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 Jul 2018 13:50
 Operator : UA/SM
 Sample : J3252-03RE
 Misc :
 ALS Vial : 141 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 TR-05-062118RE

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 05 14:10:06 2018
 Quant Method : \\TERASTORAGE\Terastorage\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ062918.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jun 29 01:53:51 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.438	3.703	400.3E6	146.7E6	9.536	10.656
2) SA Decachlor...	10.118	8.675	196.9E6	113.2E6	5.574	5.544

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ070618\
Data File : PQ030407.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 05 Jul 2018 13:50
Operator : UA/SM
Sample : J3252-03RE
Misc :
ALS Vial : 141 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
TR-05-062118RE

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 05 14:10:06 2018
Quant Method : \\TERASTORAGE\Terastorage\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ062918.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jun 29 01:53:51 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.438 min
Delta R.T.: 0.001 min
Response: 400280060
Conc: 9.54 ng/ml

Instrument :
ECD_Q
ClientSampleId :
TR-05-062118RE




#1 Tetrachloro-m-xylene

R.T.: 3.703 min
Delta R.T.: 0.000 min
Response: 146654389
Conc: 10.66 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.118 min
Delta R.T.: 0.000 min
Response: 196929647
Conc: 5.57 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.675 min
Delta R.T.: 0.000 min
Response: 113196982
Conc: 5.54 ng/ml