

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ070618\
 Data File : PQ030418.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 Jul 2018 16:45
 Operator : UA/SM
 Sample : J3816-03
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 159-WHITE-BRICK

Manual Integrations
 APPROVED

mohammad
 7/6/2018 12:29:07 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 06 02:20:25 2018
 Quant Method : \\TERASTORAGE\Terastorage\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ062918.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jun 29 01:53:51 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.438	3.700	796.0E6	253.9E6	18.965	18.448
2) SA Decachlor...	10.120	8.674	467.8E6	279.4E6	13.239m	13.684

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ070618\
Data File : PQ030418.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 05 Jul 2018 16:45
Operator : UA/SM
Sample : J3816-03
Misc :
ALS Vial : 4 Sample Multiplier: 1



Instrument :
ECD_Q
ClientSampleId :
159-WHITE-BRICK


Manual Integrations
APPROVED

mohammad
7/6/2018 12:29:07 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 06 02:20:25 2018
Quant Method : \\TERASTORAGE\Terastorage\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ062918.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jun 29 01:53:51 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.438 min
Delta R.T.: 0.001 min
Response: 796048396
Conc: 18.97 ng/ml

Instrument :
ECD_Q
ClientSampleId :
159-WHITE-BRICK


Manual Integrations
APPROVED

mohammad
7/6/2018 12:29:07 PM




#1 Tetrachloro-m-xylene

R.T.: 3.700 min
Delta R.T.: -0.003 min
Response: 253879663
Conc: 18.45 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.120 min
Delta R.T.: 0.001 min
Response: 467756141
Conc: 13.24 ng/ml m



#2 Decachlorobiphenyl

R.T.: 8.674 min
Delta R.T.: -0.001 min
Response: 279370458
Conc: 13.68 ng/ml