

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ070918\
 Data File : PQ030544.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 07 Jul 2018 14:05
 Operator : UA/SM
 Sample : J3880-09
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 SURCHARGE-SP-15

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 09 04:07:10 2018
 Quant Method : \\TERASTORAGE\Terastorage\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ062918.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jun 29 01:53:51 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.434	3.702	822.8E6	294.5E6	19.603	21.403
2) SA Decachlor...	10.111	8.670	458.8E6	259.3E6	12.986	12.701
Target Compounds						
31) L7 AR-1260-1	7.182	6.245	187.5E6	73835217	93.496	74.222
32) L7 AR-1260-2	7.431	6.584	444.0E6	74912237	187.085	65.299 #
33) L7 AR-1260-3	7.718	7.052	316.0E6	42186618	112.057	45.241 #
34) L7 AR-1260-4	8.013	7.290	149.3E6	134.7E6	74.608	55.434 #
35) L7 AR-1260-5	8.333	7.636	209.2E6	113.4E6	52.511	64.773

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ070918\
Data File : PQ030544.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Jul 2018 14:05
Operator : UA/SM
Sample : J3880-09
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
SURCHARGE-SP-15

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 09 04:07:10 2018
Quant Method : \\TERASTORAGE\Terastorage\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ062918.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jun 29 01:53:51 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

