

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\
 Method File : PQ071218.M
 Title : GC EXTRACTABLES
 Last Update : Thu Jul 12 12:41:01 2018
 Response Via : Initial Calibration

Calibration Files

500 =PQ030592.D 1000 =PQ030590.D 750 =PQ030591.D
 250 =PQ030593.D 50 =PQ030594.D

Compound		500	1000	750	250	50	Avg	%RSD
1) SA	Tetrachloro-m-xylene	4.760	4.292	4.390	4.663	4.348	4.491 E7	4.62
2) SA	Decachlorobiphenyl	3.304	3.083	3.143	3.362	3.449	3.268 E7	4.66
3) L1	AR-1016-1	6.483	5.716	5.868	6.333	6.079	6.096 E5	5.21
4) L1	AR-1016-2	9.042	8.183	8.322	8.821	8.081	8.490 E5	4.94
5) L1	AR-1016-3	1.156	1.032	1.065	1.127	1.294	1.135 E6	8.96
6) L1	AR-1016-4	9.611	8.734	8.892	9.130	8.745	9.022 E5	4.06
7) L1	AR-1016-5	1.002	0.916	0.926	1.005	1.088	0.987 E6	7.07
8) L2	AR-1221-1	5.445					5.445 E5	0.00
9) L2	AR-1221-2	3.878					3.878 E5	0.00
10) L2	AR-1221-3	1.208					1.208 E6	0.00
11) L3	AR-1232-1	6.497					6.497 E5	0.00
12) L3	AR-1232-2	3.203					3.203 E5	0.00
13) L3	AR-1232-3	4.112					4.112 E5	0.00
14) L3	AR-1232-4	3.389					3.389 E5	0.00
15) L3	AR-1232-5	3.693					3.693 E5	0.00
16) L4	AR-1242-1	7.965					7.965 E5	0.00
17) L4	AR-1242-2	1.291					1.291 E6	0.00
18) L4	AR-1242-3	1.025					1.025 E6	0.00
19) L4	AR-1242-4	9.194					9.194 E5	0.00
20) L4	AR-1242-5	9.967					9.967 E5	0.00
21) L5	AR-1248-1	1.209					1.209 E6	0.00
22) L5	AR-1248-2	1.045					1.045 E6	0.00
23) L5	AR-1248-3	1.075					1.075 E6	0.00
24) L5	AR-1248-4	1.350					1.350 E6	0.00
25) L5	AR-1248-5	1.389					1.389 E6	0.00
26) L6	AR-1254-1	1.508					1.508 E6	0.00
27) L6	AR-1254-2	2.310					2.310 E6	0.00
28) L6	AR-1254-3	2.470					2.470 E6	0.00
29) L6	AR-1254-4	1.851					1.851 E6	0.00
30) L6	AR-1254-5	1.961					1.961 E6	0.00
31) L7	AR-1260-1	1.822	1.673	1.702	1.765	1.890	1.770 E6	4.98
32) L7	AR-1260-2	2.148	2.003	2.014	2.139	2.176	2.096 E6	3.88
33) L7	AR-1260-3	2.562	2.437	2.639	2.517	2.926	2.616 E6	7.18
34) L7	AR-1260-4	1.763	1.662	1.685	1.741	1.678	1.706 E6	2.57
35) L7	AR-1260-5	3.551	3.350	3.384	3.524	3.408	3.443 E6	2.58
36) L8	AR-1262-1	1.676					1.676 E6	0.00
37) L8	AR-1262-2	1.951					1.951 E6	0.00
38) L8	AR-1262-3	2.472					2.472 E6	0.00
39) L8	AR-1262-4	2.258					2.258 E6	0.00
40) L8	AR-1262-5	4.414					4.414 E6	0.00
41) L9	AR-1268-1	4.618					4.618 E6	0.00
42) L9	AR-1268-2	4.260					4.260 E6	0.00
43) L9	AR-1268-3	3.669					3.669 E6	0.00
44) L9	AR-1268-4	1.603					1.603 E6	0.00
45) L9	AR-1268-5	1.181					1.181 E7	0.00

Signal #2 Calibration Files

500 =PQ030592.D 1000 =PQ030590.D 750 =PQ030591.D
 250 =PQ030593.D 50 =PQ030594.D

Compound		500	1000	750	250	50	Avg	%RSD
1) SA	Tetrachloro-m-xylene	1.589	1.517	1.543	1.654	1.474	1.555 E7	4.45
2) SA	Decachlorobiphenyl	2.078	1.998	2.059	2.254	2.078	2.093 E7	4.57
3) L1	AR-1016-1	3.543	3.304	3.462	3.508	3.210	3.406 E5	4.18
4) L1	AR-1016-2	3.535	3.363	3.498	3.301	3.251	3.390 E5	3.64
5) L1	AR-1016-3	2.688	2.562	2.713	2.667	3.142	2.754 E5	8.14

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\
 Method File : PQ071218.M
 Title : GC EXTRACTABLES
 Last Update : Thu Jul 12 12:41:01 2018
 Response Via : Initial Calibration

Calibration Files

500 =PQ030592.D 1000 =PQ030590.D 750 =PQ030591.D
 250 =PQ030593.D 50 =PQ030594.D

	Compound	500	1000	750	250	50	Avg		%RSD
6)	L1 AR-1016-4	3.589	3.395	3.484	3.652	3.107	3.446	E5	6.19
7)	L1 AR-1016-5	4.189	4.022	3.976	4.997	4.331	4.303	E5	9.59
8)	L2 AR-1221-1	2.286					2.286	E5	0.00
9)	L2 AR-1221-2	1.357					1.357	E5	0.00
10)	L2 AR-1221-3	4.265					4.265	E5	0.00
11)	L3 AR-1232-1	2.271					2.271	E5	0.00
12)	L3 AR-1232-2	2.006					2.006	E5	0.00
13)	L3 AR-1232-3	1.196					1.196	E5	0.00
14)	L3 AR-1232-4	1.298					1.298	E5	0.00
15)	L3 AR-1232-5	1.626					1.626	E5	0.00
16)	L4 AR-1242-1	3.221					3.221	E5	0.00
17)	L4 AR-1242-2	4.874					4.874	E5	0.00
18)	L4 AR-1242-3	2.980					2.980	E5	0.00
19)	L4 AR-1242-4	3.062					3.062	E5	0.00
20)	L4 AR-1242-5	3.483					3.483	E5	0.00
21)	L5 AR-1248-1	3.697					3.697	E5	0.00
22)	L5 AR-1248-2	3.923					3.923	E5	0.00
23)	L5 AR-1248-3	4.915					4.915	E5	0.00
24)	L5 AR-1248-4	7.668					7.668	E5	0.00
25)	L5 AR-1248-5	5.352					5.352	E5	0.00
26)	L6 AR-1254-1	8.137					8.137	E5	0.00
27)	L6 AR-1254-2	7.229					7.229	E5	0.00
28)	L6 AR-1254-3	1.226					1.226	E6	0.00
29)	L6 AR-1254-4	9.006					9.006	E5	0.00
30)	L6 AR-1254-5	4.849					4.849	E5	0.00
31)	L7 AR-1260-1	8.912	8.520	8.636	9.266	8.398	8.746	E5	3.97
32)	L7 AR-1260-2	1.038	1.009	1.021	1.113	1.003	1.037	E6	4.28
33)	L7 AR-1260-3	8.726	8.349	8.495	9.079	7.782	8.486	E5	5.66
34)	L7 AR-1260-4	2.144	2.082	2.121	2.261	2.026	2.127	E6	4.11
35)	L7 AR-1260-5	1.588	1.530	1.566	1.652	1.489	1.565	E6	3.92
36)	L8 AR-1262-1	6.869					6.869	E5	0.00
37)	L8 AR-1262-2	8.464					8.464	E5	0.00
38)	L8 AR-1262-3	1.223					1.223	E6	0.00
39)	L8 AR-1262-4	1.086					1.086	E6	0.00
40)	L8 AR-1262-5	2.341					2.341	E6	0.00
41)	L9 AR-1268-1	2.397					2.397	E6	0.00
42)	L9 AR-1268-2	2.342					2.342	E6	0.00
43)	L9 AR-1268-3	1.949					1.949	E6	0.00
44)	L9 AR-1268-4	8.090					8.090	E5	0.00
45)	L9 AR-1268-5	6.446					6.446	E6	0.00

(#) = Out of Range