

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\
 Method File : PQ071219.M
 Title : GC EXTRACTABLES
 Last Update : Sat Jul 13 02:14:45 2019
 Response Via : Initial Calibration

Calibration Files

500 =PQ041155.D 1000 =PQ041153.D 750 =PQ041154.D
 250 =PQ041156.D 50 =PQ041157.D

	Compound	500	1000	750	250	50	Avg	%RSD
1) SA	Tetrachloro-m-xylene	1.007	0.905	1.020	1.025	1.122	1.016 E7	7.60
2) SA	Decachlorobiphenyl	1.258	1.158	1.256	1.272	1.477	1.284 E7	9.11
3) L1	AR-1016-1	3.854	3.423	3.794	3.984	4.823	3.975 E5	13.02
4) L1	AR-1016-2	5.607	5.034	5.767	5.998	6.581	5.797 E5	9.74
5) L1	AR-1016-3	3.366	2.924	3.350	3.608	3.844	3.418 E5	10.01
6) L1	AR-1016-4	2.812	2.496	2.836	3.104	3.223	2.894 E5	9.79
7) L1	AR-1016-5	2.650	2.384	2.654	2.910	3.018	2.723 E5	9.12
8) L2	AR-1221-1	1.054					1.054 E5	0.00
9) L2	AR-1221-2	8.591					8.591 E4	0.00
10) L2	AR-1221-3	2.641					2.641 E5	0.00
11) L3	AR-1232-1	1.987					1.987 E5	0.00
12) L3	AR-1232-2	1.047					1.047 E5	0.00
13) L3	AR-1232-3	2.310					2.310 E5	0.00
14) L3	AR-1232-4	1.122					1.122 E5	0.00
15) L3	AR-1232-5	7.235					7.235 E4	0.00
16) L4	AR-1242-1	4.450					4.450 E5	0.00
17) L4	AR-1242-2	6.609					6.609 E5	0.00
18) L4	AR-1242-3	3.838					3.838 E5	0.00
19) L4	AR-1242-4	3.334					3.334 E5	0.00
20) L4	AR-1242-5	4.797					4.797 E5	0.00
21) L5	AR-1248-1	3.052					3.052 E5	0.00
22) L5	AR-1248-2	3.855					3.855 E5	0.00
23) L5	AR-1248-3	4.907					4.907 E5	0.00
24) L5	AR-1248-4	5.952					5.952 E5	0.00
25) L5	AR-1248-5	5.848					5.848 E5	0.00
26) L6	AR-1254-1	5.157					5.157 E5	0.00
27) L6	AR-1254-2	6.449					6.449 E5	0.00
28) L6	AR-1254-3	7.315					7.315 E5	0.00
29) L6	AR-1254-4	6.326					6.326 E5	0.00
30) L6	AR-1254-5	5.775					5.775 E5	0.00
31) L7	AR-1260-1	5.155	4.625	5.212	5.277	6.175	5.289 E5	10.57
32) L7	AR-1260-2	7.457	6.643	7.425	7.700	8.253	7.495 E5	7.75
33) L7	AR-1260-3	6.324	5.597	6.252	6.656	7.246	6.415 E5	9.39
34) L7	AR-1260-4	6.067	5.355	5.909	6.283	6.723	6.067 E5	8.28
35) L7	AR-1260-5	1.430	1.306	1.428	1.429	1.543	1.427 E6	5.87
36) L8	AR-1262-1	1.005					1.005 E6	0.00
37) L8	AR-1262-2	1.783					1.783 E6	0.00
38) L8	AR-1262-3	1.164					1.164 E6	0.00
39) L8	AR-1262-4	9.088					9.088 E5	0.00
40) L8	AR-1262-5	6.351					6.351 E5	0.00
41) L9	AR-1268-1	2.044					2.044 E6	0.00
42) L9	AR-1268-2	1.837					1.837 E6	0.00
43) L9	AR-1268-3	1.604					1.604 E6	0.00
44) L9	AR-1268-4	6.581					6.581 E5	0.00
45) L9	AR-1268-5	4.929					4.929 E6	0.00

Signal #2 Calibration Files

500 =PQ041155.D 1000 =PQ041153.D 750 =PQ041154.D
 250 =PQ041156.D 50 =PQ041157.D

	Compound	500	1000	750	250	50	Avg	%RSD
1) SA	Tetrachloro-m-xylene	6.448	6.065	6.596	6.611	6.885	6.521 E6	4.60
2) SA	Decachlorobiphenyl	4.446	4.030	4.412	4.693	5.795	4.675 E6	14.32
3) L1	AR-1016-1	2.370	2.156	2.390	2.578	3.052	2.509 E5	13.48
4) L1	AR-1016-2	3.426	3.183	3.447	3.679	4.088	3.565 E5	9.58
5) L1	AR-1016-3	1.785	1.644	1.803	1.951	2.109	1.858 E5	9.54

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\
 Method File : PQ071219.M
 Title : GC EXTRACTABLES
 Last Update : Sat Jul 13 02:14:45 2019
 Response Via : Initial Calibration

Calibration Files

500 =PQ041155.D 1000 =PQ041153.D 750 =PQ041154.D
 250 =PQ041156.D 50 =PQ041157.D

	Compound	500	1000	750	250	50	Avg		%RSD
6)	L1 AR-1016-4	1.443	1.306	1.438	1.597	1.837	1.524	E5	13.33
7)	L1 AR-1016-5	1.837	1.651	1.830	1.990	2.213	1.904	E5	11.04
8)	L2 AR-1221-1	7.408					7.408	E4	0.00
9)	L2 AR-1221-2	5.349					5.349	E4	0.00
10)	L2 AR-1221-3	1.740					1.740	E5	0.00
11)	L3 AR-1232-1	1.371					1.371	E5	0.00
12)	L3 AR-1232-2	1.456					1.456	E5	0.00
13)	L3 AR-1232-3	7.445					7.445	E4	0.00
14)	L3 AR-1232-4	6.350					6.350	E4	0.00
15)	L3 AR-1232-5	6.858					6.858	E4	0.00
16)	L4 AR-1242-1	2.876					2.876	E5	0.00
17)	L4 AR-1242-2	4.147					4.147	E5	0.00
18)	L4 AR-1242-3	2.134					2.134	E5	0.00
19)	L4 AR-1242-4	1.976					1.976	E5	0.00
20)	L4 AR-1242-5	2.970					2.970	E5	0.00
21)	L5 AR-1248-1	1.954					1.954	E5	0.00
22)	L5 AR-1248-2	2.554					2.554	E5	0.00
23)	L5 AR-1248-3	2.606					2.606	E5	0.00
24)	L5 AR-1248-4	3.173					3.173	E5	0.00
25)	L5 AR-1248-5	3.338					3.338	E5	0.00
26)	L6 AR-1254-1	3.615					3.615	E5	0.00
27)	L6 AR-1254-2	3.056					3.056	E5	0.00
28)	L6 AR-1254-3	4.958					4.958	E5	0.00
29)	L6 AR-1254-4	3.408					3.408	E5	0.00
30)	L6 AR-1254-5	4.194					4.194	E5	0.00
31)	L7 AR-1260-1	3.227	2.883	3.201	3.524	4.152	3.398	E5	14.09
32)	L7 AR-1260-2	4.199	3.757	4.217	4.565	5.498	4.447	E5	14.70
33)	L7 AR-1260-3	3.549	3.188	3.563	3.769	4.577	3.729	E5	13.89
34)	L7 AR-1260-4	2.875	2.611	2.876	3.065	3.610	3.007	E5	12.43
35)	L7 AR-1260-5	7.231	6.678	7.283	7.489	8.344	7.405	E5	8.17
36)	L8 AR-1262-1	5.434					5.434	E5	0.00
37)	L8 AR-1262-2	8.829					8.829	E5	0.00
38)	L8 AR-1262-3	3.486					3.486	E5	0.00
39)	L8 AR-1262-4	6.132					6.132	E5	0.00
40)	L8 AR-1262-5	2.619					2.619	E5	0.00
41)	L9 AR-1268-1	9.099					9.099	E5	0.00
42)	L9 AR-1268-2	8.081					8.081	E5	0.00
43)	L9 AR-1268-3	6.527					6.527	E5	0.00
44)	L9 AR-1268-4	2.754					2.754	E5	0.00
45)	L9 AR-1268-5	1.770					1.770	E6	0.00

(#) = Out of Range